


## *ABSOLUTE STANDARDS, INC.*



### **World Class Manufacturing & Internationally Recognized Quality**

As Absolute Standards enters the latter half of our second decade of service, we continue to maintain our position as the industry leader. As the largest supplier of Analytical Reference Materials (ARM)<sup>®</sup> and Performance Evaluation samples, we ship our products to all fifty states, several territories, and over thirty countries.

Focusing on the needs of the global testing body, we strive to provide the highest quality and broadest range of analytical standards to an industry that relies on accuracy and precision. Moreover, Absolute Standards is the only company in the industry that has exceeded the international requirements of both ISO/IEC 17025 and ISO 9001:2000 for the manufacture and analytical verification of our entire product line.

As an analytical standards manufacturer, we recognize our role as an important link in the analytical testing community. Your analyses require the use of accurate, traceable, and defensible reference materials. It is this requirement that drives Absolute Standards to manufacture products that lead the industry in quality and value.

Our philosophy at Absolute Standards is one of compliance. High expectations backed by quality accreditations help us achieve our goals. Discover how Absolute's commitment can help satisfy your compliance needs.

Stephen J. Arpie, *M.S. Technical Director*

John P. Criscio, *President/CEO*



---

## ORDERING & TECHNICAL INFORMATION

---

To Place an Order...  
please have the following information available:

1. Account #
2. Purchase Order #/ Credit Card #
3. Part # and Quantity

We are available for orders & technical service from 9am-6pm (EST) by Phone, and 24 hrs. a day by Fax, Mail, & Internet.

...Phone...

**800-368-1131**

...Fax...

**800-410-2577**

...Mail...

Absolute Standards, Inc.

P.O. Box 5585

Hamden, CT 06518

...Email...

customerserv@absolutestandards.com

Order on-line at **www.AbsoluteStandards.com** using LabCart™  
Secure, Simple, Searchable, & Speedy

We accept major credit cards: **MasterCard / VISA / AMEX.**

Orders that are placed before 5pm (EST) are shipped the same day by  
**FedEx & UPS.**

**Prices** are F.O.B. Hamden CT. Terms are Net 30 Days. Terms and conditions of all sales contracts will be governed by the laws of The State of Connecticut. Prices are subject to change without notice.

**Warranty** by Absolute Standards includes that our products are as described when shipped. No warranty as to fitness for any particular application is expressed or implied. Errant shipments and/or quality claims must be made within 30 days of receipt. No returns can be made without a Return Authorization Number and instructions of forwarding.

**Liability** by Absolute Standards is limited solely to the replacement of the product or refund of the purchase price. These products are intended for research use only. Owing to their hazardous nature, they should be handled only by qualified, trained personnel familiar with their use.

## CERTIFICATES & ACCREDITATIONS

Please visit our website at  
**www.AbsoluteStandards.com**  
 to download current copies of our ISO 9001, ISO 17025 and A2LA certificates.



### NSF International Strategic Registrations

789 North Dixboro Road, Ann Arbor, Michigan 48105  
 (888) NSF-9000

### Certificate of Registration

This certifies that the Quality Management System of  
**ABSOLUTE STANDARDS, INC.**

44 Rossotto Drive  
 Hamden, Connecticut, 06514, United States

has been assessed by NSF-ISR and found to be in conformance to the following standard(s):

### ISO 9001:2000

#### Scope of Registration:

Design, development, production and service of analytical reference materials (ARM), used as calibration and quality control standards.

**Exclusions:** None

#### Industrial Classification:

IAF - QMS: 12  
 SIC: 2899

Certificate Number: C0023640-IS1  
 Certificate Issue Date: 19-DEC-2008  
 Registration Date: 16-MAY-2008  
 Expiration Date \*: 15-MAY-2011

*Christian B. Lupo*

Christian B. Lupo, General Manager  
 NSF-ISR, Ltd.



Authorized Registration and/or Accreditation Marks. This certificate is property of NSF-ISR and must be returned upon request. \*Company is audited for conformance at regular intervals. To verify registrations call (888) NSF-9000 or visit our web site at [www.nsf-ir.org](http://www.nsf-ir.org)

30% recycled post-consumer fiber.



# Absolute Standards, Inc.



**Custom Standard Quotation Request Form**  
**Fax:** (800) 410-2577, Technical Service Dept, Absolute Standards, Inc.  
**Email:** customerserv@absolutestandards.com

Page \_\_\_\_\_ of \_\_\_\_\_

Date: \_\_\_\_ / \_\_\_\_ / \_\_\_\_

**From:**

Company Contact: \_\_\_\_\_

Company: \_\_\_\_\_ Account#: \_\_\_\_\_

Company Address: \_\_\_\_\_

Telephone: \_\_\_\_\_ Fax: \_\_\_\_\_

Email: \_\_\_\_\_

**Product Description:** \_\_\_\_\_

**Solvent/ Matrix:** \_\_\_\_\_

**Volume Size:**  1 x 100 mL       1 x 500 mL       Other: \_\_\_\_\_  
Qty: \_\_\_\_\_      Qty: \_\_\_\_\_      Qty: \_\_\_\_\_

Date Required: \_\_\_\_ / \_\_\_\_ / \_\_\_\_      Analysis requested? (additional charge)  Yes  No

	<u>Component</u>	<u>CAS# (optional)</u>	<u>Concentration (ug/mL)</u>
1.	_____	_____	_____
2.	_____	_____	_____
3.	_____	_____	_____
4.	_____	_____	_____
5.	_____	_____	_____
6.	_____	_____	_____
7.	_____	_____	_____
8.	_____	_____	_____
9.	_____	_____	_____
10.	_____	_____	_____
11.	_____	_____	_____
12.	_____	_____	_____
13.	_____	_____	_____
14.	_____	_____	_____
15.	_____	_____	_____
16.	_____	_____	_____
17.	_____	_____	_____
18.	_____	_____	_____
19.	_____	_____	_____
20.	_____	_____	_____

Comments: \_\_\_\_\_

# AbsoluteGrade™ - Specifications

Absolute Standards is the *only* company in the industry that has exceeded the international requirements of both ISO/IEC 17025 and ISO 9001:2000 for the manufacture and analytical verification of our *entire* product line. Equipped with state of the art instrumentation, our quality products are backed by the analytical power of an ICP-MS and four LC's. Included in this catalog are reference materials for calibration of the following instruments: AA, ICP, ICP-MS, IC, IE & several wet chemical methods. Method references are indicated wherever possible. In addition, the CLP section has been updated to include the latest revisions.

Our capabilities are diverse and expanding. As we strive for continuous improvement, we've added the following services to better accommodate your needs: bulk sealing, on-line ordering with *LabCart™*, and a fully searchable Products database available at [www.AbsoluteStandards.com](http://www.AbsoluteStandards.com). We know that your requirements can be unique and varied therefore, if you would like a custom blend, please contact us with your specifications. Let our *experienced* chemists give you a feasibility assessment of *your* design and work with you to enhance the stability and robustness of your custom formulation.



## AbsoluteGrade™ - Specifications

- Suitable for ICP AES, DCP AES, ICP MS, High-Accuracy AA Flame or Graphite Furnace work.
- Manufactured using the *Highest Purity* starting materials available (typically 99.999%).
- Digestion in acid-cleaned PTFE labware to prevent possible contamination.
- Class A volumetric glassware used for pipetting and diluting where applicable.
- Signed Certificate of Analysis included with *every* standard.
- Traceable to NIST **and other internationally recognized third party standards.**
- Formulation Uncertainty provided for every analyte
- Density measurement allows wt/wt use.
- Verification of impurities in both starting materials and final solutions.
- Immediate shipment from stock.
- Total U.S. "Right To-Know", and O.S.H.A. compliance.
- Cost-effective packaging.

## TABLE OF CONTENTS

TITLE	PAGES
Certificate of Analysis.....	7
AbsoluteGrade™ Specifications / Certified Weight Report.....	8-9
Single Component Solutions AA/ICP Spectroscopy.....	10-13
Elemental Group Kits.....	14-15
Matrix Modifiers and Blanks (AA/ICP).....	16
Graphite Furnace (GFAA) Standards.....	17
Contract Laboratory Program (CLP) Standards.....	18-35
ILM 05.2 ICP-AES.....	19-22
ILM 05.2 ICP-MS.....	23-27
ILM 05.3 ICP-AES (Water).....	28-30
ILM 05.3 ICP-AES (Soil).....	31-32
ILM 05.3 ICP-MS (Water).....	33-34
ILM 06.X ICP-MS (Soil).....	35
TCLP Multi-Element Calibration Standards.....	36-37
EPA 6010 A/B Standards.....	38-39
EPA 6020/ 6020A Standards.....	40
EPA 200.7 Standards.....	41-51
Calibration Standards.....	41
Plasma, Tuning, Lab Check Solutions.....	42
Spectral Interference.....	43
Lab Fortifying Solution (LFS).....	44
Spiking Standards.....	45-46
Drinking Water Standards.....	47
Instrument Wavelength Standards.....	49
Quality Control Standards.....	50-51
EPA 200.8 Standards.....	52-56
Tuning & Internal Standards.....	52
Calibration Standards.....	53
Verification Standards.....	54
Interference / Memory Check Solutions.....	55
Multi-Element Solutions.....	56
EPA 200.9 Standards .....	57
EPA Method 1620.....	58
EPA Method 1636.....	59
EPA Method 1638/1640.....	60
EPA Method 1664.....	61
Groundwater and Wastewater Standards (600/4-79-020).....	62-63
Trace Metals In Fish (600/4-79-020).....	63
Miscellaneous Mixes.....	64
EPA 300.0 Standards (Anions by Ion Chromatography).....	65-68
Single Components / Eluents.....	65
Multi-Analyte Solutions.....	66-67
Inorganic Disinfection By-Products.....	68

# TABLE OF CONTENTS

Cations .....	69
EPA Method 314.0/ 314.1/ 332.0 (Perchlorate).....	70
EPA Method 353.2 (Nitrate & Nitrite).....	71
Organic Acids.....	72
Miscellaneous Wet Methods.....	72-74
Part# Index.....	75-77
Custom Quotation Order Form.....	4, 78

## ABSOLUTE STANDARDS, INC.

ISO 9001 DLS ANSI-RAB ACCREDITED



### Certificate of Analysis



**Conformance:** The "Certificate of Analysis" and the "Certified Weight Report" fulfill the requirements in the current versions of: ISO Guides (9001, 24, 31, 34, 17025) & NIST-NVLAP Handbook 150-19.

**Health & Safety:** See the attached MSDS & Certified Weight Report before use.

**Intended Use:** This Analytical Reference Material (ARM<sup>®</sup>) is intended primarily for use in the characterization of unknowns & the establishment of response factors by qualified personnel. Typical instrumental organic assays include: GC & LC, and inorganic assays include: ICP & AA. The product is for laboratory use only.

**Certified Values:** In production, gravimetric/volumetric readings are certified to be within +/- 0.5% of the stated value and are valid between 18°C & 30°C. The measured uncertainty can be found on the Certified Weight Report. All product weighings are performed on an analytical balance that is calibrated to **NIST Traceable** standard weights & certified by the manufacturer. The volumetric glassware used is Class "A" type & conforms to ASTM E-288 unless otherwise stated. The solvents & compounds used are of the highest practical purity & typically meet or exceed **ACS Reagent Grade & ACS Standards Grade** specifications.

**Homogeneity & Stability:** No heterogeneity was observed in the preparation, packaging & storage of this standard. Expiration dates can be found on the Certified Weight Report.

**Purity & Identity:** Purity determinations are performed using at least one analytical technique. Identity assays are typically performed using two dissimilar methods.

Organic solutions & neat are typically formulated from materials whose purity & identity have been characterized by GC-MSD & LC-PDA techniques with comparison to a **NIST Traceable** library of mass spectra when available. Additional characterization techniques may include but are not limited to: refractive index measurements of liquids, melting point measurements of solids, & GC-FID, ECD, PID, ELCD, NPD, LC-PDA measurements for purity.


Inorganic solutions & neat are typically formulated from materials whose purity & identity have been characterized by ICPMS with comparison to a **NIST Standard Reference Material** when available. Additional characterization techniques may include but are not limited to: titrimetry, and densitometry, AA or ICP.

Final solutions are subjected to instrumental analysis to support the gravimetric values when appropriate. The data for the quality assurance testing is published with the Certified Weight Report. Typical instrumental organic assays include GC and LC. Inorganic assays include ICPMS, titrimetry, and densitometry.

**Storage:** Sealed ampules and other containers should be stored in the dark at temperatures above the freezing point of the solution and not more than 30°C. Certification by Absolute Standards, Inc. is typically valid for 5 years from the date of manufacture (Lot Number). Certified values are not applicable to opened ampules or to materials stored in re-sealable containers. See the "Certified Weight Report" for specific values.

**Usage:** Ampules & bottles should be brought to room temperature (18 to 30°C) before opening. Sonication may be required for high concentration solutions. After opening, care should be exercised to avoid concentration changes owing to evaporation of the solvent or essential components. We recommend that a suitable re-sealable container be available before opening an ampule to decant the standard for short-term storage.

**Legal Notice:** Warranty of products are as described when shipped. No warranty as to fitness for any particular application is expressed or implied. Errant shipments and/or quality claims must be made within 10 days of receipt. Liability is limited solely to the replacement of the product or refund of purchase price.

**Certifying Officer:** Stephen J. Arpie, M.S. Director 

Absolute Standards, Inc. • 44 Rossotto Drive • Hamden, CT 06514

Voice: 1-203-281-2917 • Fax: 1-203-281-2922

email: [quality@absolutestandards.com](mailto:quality@absolutestandards.com) • www: <http://www.absolutestandards.com>

Document Identification: Certificate of Analysis Rev 5, Date Issued: 01/01/2006

**CERTIFIED  
WEIGHT  
REPORT**

**AbsoluteGrade™  
SPECIFICATIONS**

All of our standards are accompanied by a *Certified Weight Report*.

**ABSOLUTE STANDARDS, INC. • 44 ROSSOTTO DRIVE • HAMDEN, CT 06514 • (800)-368-1131** ISO 9001 Quality System registered  
DLS ANSI/RAB Accredited

---

**CERTIFIED WEIGHT REPORT:**

<b>Part Number:</b> 53084	<b>Lot #</b>	<b>Solvent(s):</b>
<b>Lot Number:</b> 070102	L135898	Nitric Acid
<b>Description:</b> QC Standard 1	N37G05	Hydrofluoric acid
<b>Expiration Date:</b> 070105	2.0%	20.0
	0.025%	0.25
		(mL)
<b>Nominal Concentration (µg/mL):</b> 100	5E-05	Balance Uncertainty
	999.88	0.025
		Flask Uncertainty

*Lawrence Barry*

Formulated By: Lawrence Barry 070102

---

*Pedro L. Rentas*

Reviewed By: Pedro L. Rentas 070102

Compound	Part Number	Lot Number	Dilution Factor	Initial Volume	Uncertainty Pipette	Initial Conc. (µg/mL)	Final Conc. (µg/mL)	Expanded Uncertainty (+/-) (µg/mL)	MSDS Information (Solvent Safety Info. On Attached pg.)		
									CAS#	OSHA PEL (TWA)	LD50
1. Beryllium acetate (Be)	57104	051100	0.0100	10.00	0.006	10000.04	100.0	0.002334	19049-40-2	0.002 ug/m3	N/A
2. Calcium carbonate (Ca)	58120	012202	0.0100	10.00	0.006	10000.82	100.0	0.002334	00471-34-1	7 mg/m3	N/A
3. Cadmium nitrate (Cd)	58148	091400	0.0100	10.00	0.006	10000.19	100.0	0.002002	10022-68-1	0.2 mg/m3	N/A
4. Chromium nitrate nonahydrate (Cr)	58124	040901	0.0100	10.00	0.006	10000.34	100.0	0.002002	07789-02-8	0.5 mg/m3	ori-rat 3250mg/kg
5. Cobalt nitrate (Co)	57127	020901	0.0100	10.00	0.006	10000.54	100.0	0.002002	10026-22-9	5 mg/m3	ori-rat 694 mg/kg
6. Copper nitrate (Cu)	58129	062001	0.0100	10.00	0.006	10003.59	100.0	0.002001	03251-23-8	N/A	ori-rat 940 mg/kg
7. Iron (III) Nitrate Nonahydrate	58126	061002	0.0100	10.00	0.006	10001.80	100.0	0.002003	07782-61-8	7 mg/m3	N/A
8. Lead nitrate (Pb)	57182	032900	0.0100	10.00	0.006	9999.67	100.0	0.002002	10099-74-8	0.05 mg/m3	500 mg/kg
9. Magnesium nitrate (Mg)	58112	110601	0.0100	10.00	0.006	10000.20	100.0	0.002002	10377-60-3	7 mg/m3	N/A
10. Manganese nitrate (Mn)	57125	060800	0.0100	10.00	0.006	10000.37	100.0	0.002002	15710-66-4	5 mg/m3	N/A
11. Ammonium molybdate (Mo)	58142	082901	0.0100	10.00	0.006	10000.17	100.0	0.002002	13106-76-8	N/A	ori-rat 333 mg/kg
12. Nickel nitrate (Ni)	58128	102600	0.0100	10.00	0.006	10000.03	100.0	0.002002	10377-66-9	5 mg/m3	N/A
13. Selenium oxide (Se)	58134	011900	0.0100	10.00	0.006	10001.32	100.0	0.002002	07746-08-4	0.2 mg/m3	N/A
14. Thallium nitrate (Tl)	57181	120700	0.0100	10.00	0.006	10000.34	100.0	0.002002	10102-45-1	5 mg/m3	N/A
15. Ammonium hexafluorotellurate (Tl)	57122	042400	0.0100	10.00	0.006	9999.90	100.0	0.002002	16962-40-6	N/A	N/A
16. Ammonium Metavanadate (V)	58123	110701	0.0100	10.00	0.006	10000.32	100.0	0.002002	07803-55-6	1.0 mg/m3	ori-rat 630 mg/kg
17. Zinc nitrate (Zn)	58130	041101	0.0100	10.00	0.006	10000.21	100.0	0.002002	10196-18-6	1 mg/m3	ori-rat 1190mg/kg
18. Antimony Oxide (Sb)	58151	071001	0.0100	10.00	0.006	10000.04	100.0	0.002002	07440-36-0	5.0 mg/m3	N/A
19. Arsenic (As)	58133	062000	0.0100	10.00	0.006	10000.48	100.0	0.002002	07440-38-2	0.2 mg/m3	N/A

[1] Spectrum No.1 | 17.524 sec|S3084.D# [Count] [Linear]


**ASSAYED ON AN  
HP 4500 ICP MS**



# AbsoluteGrade™ SPECIFICATIONS

## CERTIFIED WEIGHT REPORT

A Certified Weight Report is generated for each lot illustrating the formulation procedure, traceability, & actual analysis.

<b>Absolute Standards, Inc.</b>		ISO 9001 Quality System Registered																	
44 Rossotto Drive • Hamden, CT • 06514 • (800) 368-1131		DLS ANSI-RAB Accredited																	
<b>AbsoluteGrade™ Solution</b>			Part# 53084																
<b>Certified Concentration (µg/mL)</b>	(µg/mL)	(+, - µg/mL)	Lot# 070102																
	100.0	0.002																	
The certified value is the concentration calculated from gravimetric and volumetric measurements unless otherwise specified.																			
<b>TRACEABILITY DOCUMENTATION:</b>																			
<b>A) Classical Chemical Analysis:</b>																			
<u>Method</u>	<u>Traceability</u>	<u>Concentration</u>																	
EDTA Titration	NIST SRM 928 Lead Nitrate	N/A																	
Gravimetric Analysis	NIST Weights	N/A																	
Redox Titration	NIST SRM 136e Potassium Dichromate	N/A																	
Volhard Titration	NIST SRM 999 Potassium Chloride	N/A																	
<b>B) Instrumental Analysis by Inductively Coupled Plasma Mass Spectroscopy (ICP-MS):</b>																			
This analytical solution was tested against the appropriate NIST SRM's:																			
<b>Trace Metals Verification by ICP-MS (µg/mL)</b>																			
Al	<0.02	Cd	*	Dy	<0.02	Hf	<0.02	Li	<0.02	Ni	*	Pr	<0.02	Se	*	Tb	<0.02	W	<0.02
Sb	*	Ca	*	Er	<0.02	Ho	<0.02	Lu	<0.02	Nb	<0.02	Re	<0.02	Si	<0.02	Te	<0.02	U	<0.02
As	*	Ce	<0.02	Eu	<0.02	In	<0.02	Mg	*	Os	<0.02	Rh	<0.02	Ag	<0.02	Tl	*	V	*
Ba	<0.02	Cs	<0.02	Gd	<0.02	Ir	<0.02	Mn	*	Pd	<0.02	Rb	<0.02	Na	<0.2	Th	<0.02	Yb	<0.02
Be	*	Cr	*	Ga	<0.02	Fe	*	Hg	<0.2	P	<0.02	Ru	<0.02	Sr	<0.02	Tm	<0.02	Y	<0.02
Bi	<0.02	Co	*	Ge	<0.02	La	<0.02	Mo	*	Pt	<0.02	Sm	<0.02	S	<0.02	Sn	<0.02	Zn	*
B	<0.02	Cu	*	Au	<0.02	Pb	*	Nd	<0.02	K	<0.2	Sc	<0.02	Ta	<0.02	Ti	*	Zr	<0.02
(*) =Target Element																			
<b>C) Physical Characterization:</b>																			
Analyzed Density of Solution (g/mL):		1.018		<b>Certified by:</b>															
Temperature (°C):		22.6																	
Homogeneity: No heterogeneity was observed in the preparation of this standard.																			
<p>We use purified acids, 18 meg ohm double deionized water, calibrated Class A glassware and the highest purity raw materials available, (typically 99.999%). We meticulously clean our bottles by acid leaching and then triple rinsing with ASTM Type I water prior to use. Our standards are made gravimetrically using balances that are calibrated with weights traceable to NIST. (NIST Test #: 732/245790). We certify that all our standards are within the stated uncertainties unless otherwise stated, assuming that the bottle is kept tightly capped and stored under normal laboratory conditions.</p>																			
<p>Reference: Taylor, B.N. and Kuyatt, C.E., "Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Result," NIST Technical Note 1297, U.S. government Printing Office, Washington DC, (1994).</p>																			
P# 53084 L#070102, Report		Page 2 of 2		Printed on 10/05/2002, 11:26AM															

**SINGLE  
COMPONENT  
SOLUTIONS**
**PLASMA EMISSION SPECTROSCOPY  
REFERENCE STANDARDS**
**1,000 ug/mL Single Components for ICP/AA**

Element		Matrix	Part#	\$/100mL	Part#	\$/500mL
Aluminum	Al	Al(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57013	30	58013	70
Antimony	Sb	Sb <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub> tr.Tartaric acid	57051	30	58051	70
Arsenic	As	As <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57033	30	58033	70
Barium	Ba	Ba(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57056	30	58056	70
Beryllium	Be	Be <sub>4</sub> O(C <sub>2</sub> H <sub>3</sub> O <sub>2</sub> ) <sub>6</sub> /HNO <sub>3</sub>	57004	30	58004	70
Bismuth	Bi	Bi(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57083	30	58083	70
Boron	B	H <sub>3</sub> BO <sub>3</sub> /H <sub>2</sub> O	57005	30	58005	70
Cadmium	Cd	Cd(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57048	30	58048	70
Calcium	Ca	CaCO <sub>3</sub> /HNO <sub>3</sub>	57020	30	58020	70
Carbon	C	C(Citric Acid)/HNO <sub>3</sub>	57006	30	58006	70
Cerium	Ce	Ce(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57058	30	58058	70
Cesium	Cs	CsNO <sub>3</sub> /HNO <sub>3</sub>	57055	30	58055	70
Chromium	Cr	Cr(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57024	30	58024	70
Chromium-Cr <sup>6+</sup>	Cr <sup>6+</sup>	(NH <sub>4</sub> ) <sub>2</sub> Cr <sub>2</sub> O <sub>7</sub> /H <sub>2</sub> O	54161	30	54172	50
Cobalt	Co	Co(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57027	30	58027	70
Copper	Cu	Cu(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57029	30	58029	70
Dysprosium	Dy	Dy <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57066	30	58066	75
Erbium	Er	Er <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57068	30	58068	75
Europium	Eu	Eu <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57063	30	58063	75
Gadolinium	Gd	Gd <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57064	30	58064	75
Gallium	Ga	Ga <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57031	75	58031	150
Germanium	Ge	(NH <sub>4</sub> ) <sub>2</sub> GeF <sub>6</sub> /tr, HF	57032	30	58032	70
Gold	Au	NH <sub>4</sub> AuCl <sub>4</sub> /HCl	57079	75	58079	150
Hafnium	Hf	Hf <sub>2</sub> O <sub>3</sub> /HCl	57072	30	58072	75
Holmium	Ho	Ho <sub>3</sub> O <sub>3</sub> /HNO <sub>3</sub>	57067	30	58067	75
Indium	In	In <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57049	30	58049	70
Iridium	Ir	IrCl <sub>3</sub> /HCl	57077	75	58077	150
Iron	Fe	Fe(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57026	30	58026	70
Iron-Ferrous	Fe <sup>2+</sup>	(NH <sub>4</sub> )Fe <sup>2+</sup> (SO <sub>4</sub> ) <sub>2</sub> /H <sub>2</sub> SO <sub>4</sub>	54141	30	54174	50
Iron-Total	Fe	[Fe <sup>2+</sup> ]+[Fe <sup>3+</sup> ]/H <sub>2</sub> SO <sub>4</sub>	54140	30	54173	50
Lanthanum	La	LaCl/HNO <sub>3</sub>	57057	30	58057	70
Lead	Pb	Pb(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57082	30	58082	70
Lithium	Li	LiNO <sub>3</sub> /HNO <sub>3</sub>	57003	30	58003	70
Lithium 6 <sup>+</sup>		Li <sup>6+</sup> NO <sub>3</sub> /HNO <sub>3</sub>	59021	200	59097	1000
Lutetium	Lu	Lu <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57071	30	58071	75
Magnesium	Mg	Mg(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57012	30	58012	70
Manganese	Mn	Mn(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57025	30	58025	70
Mercury	Hg	Hg(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57080	30	58080	70
Mercury-Organic	Hg	MeHgCl/HNO <sub>3</sub>	54170	30	54171	50
Mercury-Total	Hg	Hg(NO <sub>3</sub> ) <sub>2</sub> +MeHgCl/HNO <sub>3</sub>	54004	30	54168	50

## PLASMA EMISSION SPECTROSCOPY REFERENCE STANDARDS

## SINGLE COMPONENT SOLUTIONS

### 1,000 ug/mL Single Components for ICP/AA

Element		Matrix	Part#	\$/100mL	Part#	\$/500mL
Molybdenum	Mo	(NH <sub>4</sub> )MoO <sub>4</sub> /H <sub>2</sub> O	57042	30	58042	70
Neodymium	Nd	Nd <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57060	30	58060	75
Nickel	Ni	Ni(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57028	30	58028	70
Niobium	Nb	(NH <sub>4</sub> )NbF <sub>6</sub> /tr.HF	57041	30	58041	70
Palladium	Pd	Pd/HNO <sub>3</sub>	57046	125	58046	200
Phosphorus	P	(NH <sub>4</sub> )H <sub>2</sub> PO <sub>4</sub> /HNO <sub>3</sub>	57015	30	58015	70
Platinum	Pt	[Pt(NH <sub>3</sub> ) <sub>4</sub> ](NO <sub>3</sub> ) <sub>2</sub> /HCL	57078	75	58078	150
Potassium	K	KNO <sub>3</sub> /HNO <sub>3</sub>	57019	30	58019	70
Praesodymium	Pr	Pr <sub>6</sub> O <sub>11</sub> /HNO <sub>3</sub>	57059	30	58059	75
Rhenium	Re	Re/HNO <sub>3</sub>	57075	125	58075	150
Rhodium	Rh	RhCl <sub>3</sub> /HCl	57045	100	58045	400
Rubidium	Rb	RbNO <sub>3</sub> /HNO <sub>3</sub>	57037	30	58037	75
Ruthenium	Ru	RuCl <sub>3</sub> /HNO <sub>3</sub>	57044	75	58044	150
Samarium	Sm	Sm <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57062	30	58062	75
Scandium	Sc	Sc(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57021	75	58021	150
Selenium	Se	SeO <sub>2</sub> /HNO <sub>3</sub>	57034	30	58034	70
Silica	SiO <sub>2</sub>	SiO <sub>2</sub> /NaOH	54159	30	54169	75
Silicon	Si	(NH <sub>4</sub> ) <sub>2</sub> SiF <sub>6</sub> /HNO <sub>3</sub>	57014	30	58014	70
Silver	Ag	AgNO <sub>3</sub> /HNO <sub>3</sub>	57047	30	58047	70
Sodium	Na	NaNO <sub>3</sub> /HNO <sub>3</sub>	57011	30	58011	70
Strontium	Sr	Sr(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57038	30	58038	70
Sulfur	S	(NH <sub>4</sub> ) <sub>2</sub> SO <sub>4</sub> /H <sub>2</sub> O	57016	30	58016	70
Tantalum	Ta	NH <sub>4</sub> TaF <sub>6</sub> /tr.HF	57073	30	58073	75
Tellurium	Te	TeO <sub>2</sub> /HCl	57052	50	58073	150
Terbium	Tb	Tb <sub>4</sub> O <sub>7</sub> /HNO <sub>3</sub>	57065	30	58065	75
Thallium	Tl	TlNO <sub>3</sub> /HNO <sub>3</sub>	57081	30	58081	70
Thorium	Th	Th(NO <sub>3</sub> ) <sub>4</sub> /HNO <sub>3</sub>	57090	30	58090	70
Thulium	Tm	Tm <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57069	30	58069	75
Tin	Sn	(NH <sub>4</sub> ) <sub>2</sub> SnF <sub>6</sub> /HNO <sub>3</sub> /HCL	57050	30	58050	70
Titanium	Ti	(NH <sub>4</sub> ) <sub>2</sub> TiF <sub>6</sub> /HNO <sub>3</sub> /tr.HF	57022	30	58022	70
Tungsten	W	(NH <sub>4</sub> ) <sub>2</sub> WO <sub>4</sub> /H <sub>2</sub> O	57074	30	58074	70
Uranium	U	U <sub>3</sub> O <sub>8</sub> /HNO <sub>3</sub>	57092	30	58092	70
Vanadium	V	NH <sub>4</sub> VO <sub>3</sub> /HNO <sub>3</sub>	57023	30	58023	70
Ytterbium	Yb	Yb <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57070	30	58070	75
Yttrium	Y	Y <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57039	30	58039	70
Zinc	Zn	Zn(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57030	30	58030	70
Zirconium	Zr	ZrO <sub>2</sub> /HNO <sub>3</sub>	57040	30	58040	70

*For Additional Concentrations & Matrices Visit*

**[www.AbsoluteStandards.com](http://www.AbsoluteStandards.com)**

*or Call our Technical Department*

**Call Toll-Free 800-368-1131**

**SINGLE  
COMPONENT  
SOLUTIONS**
**PLASMA EMISSION SPECTROSCOPY  
REFERENCE STANDARDS**
**10,000 ug/mL Single Components for ICP/AA**

Element		Matrix	Part#	\$/100mL	Part#	\$/500mL
Aluminum	Al	Al(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57113	85	58113	160
Antimony	Sb	Sb <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub> tr.Tartaric acid	57151	85	58151	160
Arsenic	As	As <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57133	85	58133	160
Barium	Ba	Ba(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57156	85	58156	160
Beryllium	Be	Be <sub>4</sub> O(C <sub>2</sub> H <sub>3</sub> O <sub>2</sub> ) <sub>6</sub> /HNO <sub>3</sub>	57104	85	58104	160
Bismuth	Bi	Bi(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57183	85	58183	160
Boron	B	H <sub>3</sub> BO <sub>3</sub> /H <sub>2</sub> O	57105	85	58105	160
Cadmium	Cd	Cd(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57148	85	58148	160
Calcium	Ca	CaCO <sub>3</sub> /HNO <sub>3</sub>	57120	85	58120	160
Carbon	C	C(Citric Acid)/HNO <sub>3</sub>	57106	85	58106	160
Cerium	Ce	Ce(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Cesium	Cs	CsNO <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Chromium	Cr	Cr(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57124	85	58124	160
Chromium-Cr <sup>6+</sup>	Cr <sup>6+</sup>	(NH <sub>4</sub> ) <sub>2</sub> Cr <sub>2</sub> O <sub>7</sub> /H <sub>2</sub> O	54175	85	54176	160
Cobalt	Co	Co(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57127	85	58127	160
Copper	Cu	Cu(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57129	85	58129	160
Dysprosium	Dy	Dy <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Erbium	Er	Er <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Europium	Eu	Eu <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Gadolinium	Gd	Gd <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Gallium	Ga	Ga <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Germanium	Ge	(NH <sub>4</sub> ) <sub>2</sub> GeF <sub>6</sub> /tr, HF	NA	NA	NA	NA
Gold	Au	NH <sub>4</sub> AuCl <sub>4</sub> /HCl	57179	300	58179	600
Hafnium	Hf	Hf <sub>2</sub> O <sub>3</sub> /HCl	NA	NA	NA	NA
Holmium	Ho	Ho <sub>3</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Indium	In	In <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Iridium	Ir	IrCl <sub>3</sub> /HCl	NA	NA	NA	NA
Iron	Fe	Fe(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57126	85	58126	160
Iron-Ferrous	Fe <sup>2+</sup>	(NH <sub>4</sub> )Fe <sup>2+</sup> (SO <sub>4</sub> ) <sub>2</sub> /H <sub>2</sub> SO <sub>4</sub>	NA	NA	NA	NA
Iron-Total	Fe	[Fe <sup>2+</sup> ]+[Fe <sup>3+</sup> ]/H <sub>2</sub> SO <sub>4</sub>	NA	NA	NA	NA
Lanthanum	La	LaCl/HNO <sub>3</sub>	NA	NA	NA	NA
Lead	Pb	Pb(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57182	85	58182	160
Lithium	Li	LiNO <sub>3</sub> /HNO <sub>3</sub>	57103	85	58103	160
Lithium 6 <sup>+</sup>	Li <sup>6+</sup>	Li <sup>6+</sup> NO <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Lutetium	Lu	Lu <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Magnesium	Mg	Mg(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57112	85	58112	160
Manganese	Mn	Mn(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57125	85	58125	160
Mercury	Hg	Hg(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57180	85	58180	160
Mercury-Organic	Hg	MeHgCl/HNO <sub>3</sub>	NA	NA	NA	NA
Mercury-Total	Hg	Hg(NO <sub>3</sub> ) <sub>2</sub> +MeHgCl/HNO <sub>3</sub>	54178	100	NA	NA

## PLASMA EMISSION SPECTROSCOPY REFERENCE STANDARDS

## SINGLE COMPONENT SOLUTIONS

### 10,000 ug/mL Single Components for ICP/AA

Element		Matrix	Part#	\$/100mL	Part#	\$/500mL
Molybdenum	Mo	(NH <sub>4</sub> )MoO <sub>4</sub> /H <sub>2</sub> O	57142	85	58142	160
Neodymium	Nd	Nd <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Nickel	Ni	Ni(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57128	85	58128	160
Niobium	Nb	(NH <sub>4</sub> )NbF <sub>6</sub> /tr.HF	NA	NA	NA	NA
Palladium	Pd	Pd/HNO <sub>3</sub>	57146	500	58146	900
Phosphorus	P	(NH <sub>4</sub> )H <sub>2</sub> PO <sub>4</sub> /HNO <sub>3</sub>	57115	85	58115	160
Platinum	Pt	[Pt(NH <sub>3</sub> ) <sub>4</sub> ](NO <sub>3</sub> ) <sub>2</sub> /HCL	57178	300	58178	600
Potassium	K	KNO <sub>3</sub> /HNO <sub>3</sub>	57119	85	58119	160
Praesodymium	Pr	Pr <sub>6</sub> O <sub>11</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Rhenium	Re	Re/HNO <sub>3</sub>	NA	NA	NA	NA
Rhodium	Rh	RhCl <sub>3</sub> /HCl	NA	NA	NA	NA
Rubidium	Rb	RbNO <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Ruthenium	Ru	RuCl <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Samarium	Sm	Sm <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Scandium	Sc	Sc(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Selenium	Se	SeO <sub>2</sub> /HNO <sub>3</sub>	57134	85	58134	160
Silica	SiO <sub>2</sub>	SiO <sub>2</sub> /NaOH	NA	NA	NA	NA
Silicon	Si	(NH <sub>4</sub> ) <sub>2</sub> SiF <sub>6</sub> /HNO <sub>3</sub>	57114	85	58114	160
Silver	Ag	AgNO <sub>3</sub> /HNO <sub>3</sub>	57147	85	58147	160
Sodium	Na	NaNO <sub>3</sub> /HNO <sub>3</sub>	57111	85	58111	160
Strontium	Sr	Sr(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57138	85	58138	160
Sulfur	S	(NH <sub>4</sub> ) <sub>2</sub> SO <sub>4</sub> /H <sub>2</sub> O	57116	85	58116	160
Tantalum	Ta	NH <sub>4</sub> TaF <sub>6</sub> /tr.HF	NA	NA	NA	NA
Tellurium	Te	TeO <sub>2</sub> /HCl	NA	NA	NA	NA
Terbium	Tb	Tb <sub>4</sub> O <sub>7</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Thallium	Tl	TlNO <sub>3</sub> /HNO <sub>3</sub>	57181	85	58181	160
Thorium	Th	Th(NO <sub>3</sub> ) <sub>4</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Thulium	Tm	Tm <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Tin	Sn	(NH <sub>4</sub> ) <sub>2</sub> SnF <sub>6</sub> /HNO <sub>3</sub> /HCL	57150	85	58150	160
Titanium	Ti	(NH <sub>4</sub> ) <sub>2</sub> TiF <sub>6</sub> /HNO <sub>3</sub> /tr.HF	57122	85	58122	160
Tungsten	W	(NH <sub>4</sub> ) <sub>2</sub> WO <sub>4</sub> /H <sub>2</sub> O	57174	85	58174	160
Uranium	U	U <sub>3</sub> O <sub>8</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Vanadium	V	NH <sub>4</sub> VO <sub>3</sub> /HNO <sub>3</sub>	57123	85	58123	160
Ytterbium	Yb	Yb <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	NA	NA	NA	NA
Yttrium	Y	Y <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57139	85	58139	160
Zinc	Zn	Zn(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57130	85	58130	160
Zirconium	Zr	ZrO <sub>2</sub> /HNO <sub>3</sub>	57140	85	58140	160

*For Additional Concentrations & Matrices Visit*

**[www.AbsoluteStandards.com](http://www.AbsoluteStandards.com)**

*or Call our Technical Department*

**Call Toll-Free 800-368-1131**

## KITS

## ELEMENTAL GROUP KITS

**Group I Elements Starter Kit**  
 (5 x 100ml bottles @ 1000ug/mL Each)

Element	Matrix
Lithium Li	LiNO <sub>3</sub> /HNO <sub>3</sub>
Sodium Na	NaNO <sub>3</sub> /HNO <sub>3</sub>
Potassium K	KNO <sub>3</sub> /HNO <sub>3</sub>
Rubidium Rb	RbNO <sub>3</sub> /HNO <sub>3</sub>
Cesium Cs	CsNO <sub>3</sub> /HNO <sub>3</sub>

Part # 52253 \$80/ 5 x 100 mL

**Group II Elements Starter Kit**  
 (5 x 100ml bottles @ 1000ug/mL Each)

Element	Matrix
Beryllium Be	Be <sub>4</sub> O(C <sub>2</sub> H <sub>3</sub> O <sub>2</sub> ) <sub>6</sub> /HNO <sub>3</sub>
Magnesium Mg	Mg(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Calcium Ca	CaCO <sub>3</sub> /HNO <sub>3</sub>
Strontium Sr	Sr(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Barium Ba	Ba(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>

Part # 52255 \$80/ 5 x 100 mL

**Transition Metals Starter Kit**  
 (9 x 100ml bottles @ 1000ug/mL Each)

Element	Matrix
Titanium Ti	(NH <sub>4</sub> ) <sub>2</sub> TiF <sub>6</sub> /HNO <sub>3</sub> /tr.HF
Vanadium V	NH <sub>4</sub> VO <sub>3</sub> /HNO <sub>3</sub>
Chromium Cr	Cr(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>
Manganese Mn	Mn(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Iron Fe	Fe(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>
Cobalt Co	Co(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Nickel Ni	Ni(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Copper Cu	Cu(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Zinc Zn	Zn(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>

Part # 52256 \$150/ 9 x 100 mL

**Lanthanide Metals Starter Kit**  
 (14 x 100ml bottles @ 1000ug/mL Each)

Element	Matrix
Lanthanum La	LaCl/HNO <sub>3</sub>
Cerium Ce	Ce(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>
Praesodymium Pr	Pr <sub>6</sub> O <sub>11</sub> /HNO <sub>3</sub>
Neodymium Nd	Nd <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>
Samarium Sm	Sm <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>
Europium Eu	Eu <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>
Gadolinium Gd	Gd <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>
Terbium Tb	Tb <sub>4</sub> O <sub>7</sub> /HNO <sub>3</sub>
Dysprosium Dy	Dy <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>
Holmium Ho	Ho <sub>3</sub> O <sub>3</sub> /HNO <sub>3</sub>
Erbium Er	Er <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>
Thulium Tm	Tm <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>
Ytterbium Yb	Yb <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>
Lutetium Lu	Lu <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>

Part # 52254 \$240/ 14 x 100 mL

## METHOD SPECIFIC SINGLE ELEMENT KITS

## KITS

## Method 200.7

Complete Single Element Kit  
(32 x 100ml bottles @ 1000ug/mL Each)

Element		Matrix
Aluminum	Al	Al(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>
Antimony	Sb	Sb <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub> tr.Tartaric acid
Arsenic	As	As <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>
Barium	Ba	Ba(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Beryllium	Be	Be <sub>4</sub> O(C <sub>2</sub> H <sub>3</sub> O <sub>2</sub> ) <sub>6</sub> /HNO <sub>3</sub>
Boron	B	H <sub>3</sub> BO <sub>3</sub> /H <sub>2</sub> O
Cadmium	Cd	Cd(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Calcium	Ca	CaCO <sub>3</sub> /HNO <sub>3</sub>
Cerium	Ce	Ce(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>
Chromium	Cr	Cr(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>
Cobalt	Co	Co(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Copper	Cu	Cu(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Iron	Fe	Fe(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>
Lead	Pb	Pb(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Lithium	Li	LiNO <sub>3</sub> /HNO <sub>3</sub>
Magnesium	Mg	Mg(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Manganese	Mn	Mn(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Mercury	Hg	Hg(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Molybdenum	Mo	(NH <sub>4</sub> ) <sub>2</sub> MoO <sub>4</sub> /H <sub>2</sub> O
Nickel	Ni	Ni(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Phosphorus	P	(NH <sub>4</sub> ) <sub>2</sub> H <sub>2</sub> PO <sub>4</sub> /HNO <sub>3</sub>
Potassium	K	KNO <sub>3</sub> /HNO <sub>3</sub>
Selenium	Se	SeO <sub>2</sub> /HNO <sub>3</sub>
Silica	SiO <sub>2</sub>	SiO <sub>2</sub> /NaOH
Silver	Ag	AgNO <sub>3</sub> /HNO <sub>3</sub>
Sodium	Na	NaNO <sub>3</sub> /HNO <sub>3</sub>
Strontium	Sr	Sr(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Thallium	Tl	TlNO <sub>3</sub> /HNO <sub>3</sub>
Tin	Sn	(NH <sub>4</sub> ) <sub>2</sub> SnF <sub>6</sub> /HNO <sub>3</sub> /HCl
Titanium	Ti	(NH <sub>4</sub> ) <sub>2</sub> TiF <sub>6</sub> /HNO <sub>3</sub> /tr.HF
Vanadium	V	NH <sub>4</sub> VO <sub>3</sub> /HNO <sub>3</sub>
Zinc	Zn	Zn(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>

Part # 52250 \$600/ 32 x 100 mL

Search our complete  
Products Database on-line at  
[www.AbsoluteStandards.com](http://www.AbsoluteStandards.com)


Search by: Component  
# of Components  
Concentration  
EPA Method

## Method 200.8

Complete Single Element Kit  
(21 x 100ml bottles @ 1000ug/mL Each)

Element		Matrix
Aluminum	Al	Al(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>
Antimony	Sb	Sb <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub> tr.Tartaric acid
Arsenic	As	As <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>
Barium	Ba	Ba(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Beryllium	Be	Be <sub>4</sub> O(C <sub>2</sub> H <sub>3</sub> O <sub>2</sub> ) <sub>6</sub> /HNO <sub>3</sub>
Cadmium	Cd	Cd(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Chromium	Cr	Cr(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>
Cobalt	Co	Co(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Copper	Cu	Cu(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Lead	Pb	Pb(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Manganese	Mn	Mn(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Mercury	Hg	Hg(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Molybdenum	Mo	(NH <sub>4</sub> ) <sub>2</sub> MoO <sub>4</sub> /H <sub>2</sub> O
Nickel	Ni	Ni(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>
Selenium	Se	SeO <sub>2</sub> /HNO <sub>3</sub>
Silver	Ag	AgNO <sub>3</sub> /HNO <sub>3</sub>
Thallium	Tl	TlNO <sub>3</sub> /HNO <sub>3</sub>
Thorium	Th	Th(NO <sub>3</sub> ) <sub>4</sub> /HNO <sub>3</sub>
Uranium	U	U <sub>3</sub> O <sub>8</sub> /HNO <sub>3</sub>
Vanadium	V	NH <sub>4</sub> VO <sub>3</sub> /HNO <sub>3</sub>
Zinc	Zn	Zn(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>

Part # 52251 \$380/ 21 x 100 mL



## MATRIX BLANKS & MODIFIERS

## CALIBRATION BLANKS AND MATRIX MODIFIERS

### MATRIX BLANKS

Blank solutions are prepared with High-Purity Acids and ASTM Type I Water. Blanks can be used to dilute your multi-element standards to run directly to establish your baseline. An Aqua Regia Blank can be prepared by mixing one part Nitric Acid Blank with three parts HCl Acid Blank.

#### Water Blank

ASTM Type I H<sub>2</sub>O

Part # 52000     \$25/ 100mL  
Part # 53000     \$40/ 500mL

#### Nitric Acid Blank

Matrix 5% HNO<sub>3</sub>

Part # 52002     \$25/100mL  
Part # 53002     \$40/500mL

#### HCl Acid Blank

Matrix 5% HCl

Part # 52001     \$25/ 100mL  
Part # 53001     \$40/ 500mL

### GRAPHITE FURNACE ATOMIC ABSORPTION MATRIX MODIFIERS

Matrix Modifiers suppress background interferences and help enhance sensitivity by changing the behavior of the element or the matrix in solution with respect to temperature. These modifiers help prevent the loss of certain elements during pyrolysis by converting the target to a less volatile form. This allows a more optimum graphite furnace program to be utilized.

#### Palladium and Magnesium Nitrate Modifier

Starting Material Pd & Mg(NO<sub>3</sub>)<sub>2</sub>

Matrix 0.3%(3 mg/mL) Pd & 0.2%(2 mg/mL) Mg in 5% HNO<sub>3</sub>

Part # 52311     \$250/100mL  
Part # 53311     \$450/500mL

#### Palladium Nitrate Modifier

Starting Material Pd

Matrix 1%(10 mg/mL) in 2% HNO<sub>3</sub>

Part # 57146     \$500/100mL  
Part # 58146     \$900/500mL

#### Nickel Nitrate Modifier

Starting Material Ni(NO<sub>3</sub>)<sub>2</sub>

Matrix 5%(50 mg/mL) in 5% HNO<sub>3</sub>

Part # 52003     \$250/100mL  
Part # 53003     \$350/500mL

#### Ammonium Dihydrogen Phosphate Modifier

Starting Material (NH<sub>4</sub>)H<sub>2</sub>PO<sub>4</sub>

Matrix 10%(100 mg/mL) in H<sub>2</sub>O

Part # 52005     \$350/100mL  
Part # 53005     \$500/500mL

#### Lanthanum Nitrate Modifier

Starting Material LaNO<sub>3</sub>

Matrix 5%(50 mg/mL) in 5% HNO<sub>3</sub>

Part # 52312     \$250/100mL  
Part # 53312     \$350/500mL

#### Magnesium Nitrate Modifier

Starting Material Mg(NO<sub>3</sub>)<sub>2</sub>

Matrix 2%(20 mg/mL) in 5% HNO<sub>3</sub>

Part # 52004     \$125/100mL  
Part # 53004     \$200/500mL

#### Ammonium Nitrate Modifier

Starting Material NH<sub>4</sub>NO<sub>3</sub>

Matrix 2%(20 mg/mL) in H<sub>2</sub>O

Part # 52309     \$250/100mL  
Part # 53309     \$350/500mL

#### Calcium Nitrate Modifier

Starting Material Ca(NO<sub>3</sub>)<sub>2</sub> · 4H<sub>2</sub>O

Matrix 2%(20 mg/mL) in 5% HNO<sub>3</sub>

Part # 52310     \$250/100mL  
Part # 53310     \$350/500mL

#### Lanthanum Chloride Modifier

Starting Material LaCl<sub>3</sub>

Matrix 5%(50 mg/mL) in 5% HCl

Part # 52313     \$250/100mL  
Part # 53313     \$350/500mL



## GFAA CALIBRATION AND SPIKING SOLUTIONS

## GFAA STANDARDS

### GRAPHITE FURNACE AA CALIBRATION AND SPIKING

#### GFAA Calibration

6 Components	
Matrix 5% HNO <sub>3</sub>	
Element	Conc.(ug/mL)
Sb	100
As	50
Cd	10
Pb	50
Se	100
Tl	50

**Part # 52314**      **\$85/100 mL**  
**Part # 53314**      **\$190/500 mL**

#### GFAA Initial Calibration

##### Verification (2nd Source)

6 Components	
Matrix 5% HNO <sub>3</sub>	
Element	Conc.(ug/mL)
Sb	50
As	25
Cd	5
Pb	25
Se	50
Tl	25

**Part # 52315**      **\$85/100 mL**  
**Part # 53315**      **\$190/500 mL**

#### GFAA Spike Standard

6 Components	
Matrix 5% HNO <sub>3</sub>	
Element	Conc.(ug/mL)
Sb	100
As	40
Cd	5
Pb	20
Se	10
Tl	50

**Part # 52316**      **\$85/100 mL**  
**Part # 53316**      **\$190/500 mL**

#### GFAA Mercury Standard

Matrix 2% HNO <sub>3</sub>	
Element	Conc.(ug/mL)
Hg	100

**Part # 57280**      **\$30/100 mL**  
**Part # 58280**      **\$60/500 mL**

**CLP****USEPA CONTRACT LABORATORY PROGRAM  
STATEMENT OF WORK FOR INORGANIC ANALYSIS**

The following section illustrates the solutions appropriate for the inorganic analytical methods derived from the USEPA Contract Laboratory Program. These methods are specific for ICP-AES and ICP-MS technologies and are suitable for the detection of metals in various matrices. These methods are designed to not only isolate, but to both qualitatively and quantitatively detect specific target metals in water and soil/sediment.

**ACRONYM    DESCRIPTION**

<b>Cal</b>	Calibration Standards
<b>CCV</b>	Continuing Calibration Verification: To ensure calibration accuracy during each analysis, run a CCV standard for every wavelength/ mass used for each analyte.
<b>CRQL</b>	Contract Required Quantation Limit
<b>CRI</b>	Standard prepared at the CRQL level
<b>ICB/CCB</b>	Initial and Continuing Calibration Blank
<b>ICS</b>	Interference Check Sample: To verify inter-element and background correction factors, the Contractor shall analyze and report the results for the ICS, for all elements on the Target Analyte List (TAL), and for all interferents (target and non-target), at the beginning and end of each analysis run, but not before the ICV. In addition, the Contractor shall analyze and report the results for the ICS at a frequency of not less than once per 20 analytical samples per analysis run.
<b>ICV</b>	Initial Calibration/Verification: Immediately after each of system has been calibrated, the accuracy of the initial calibration shall be verified and documented for every analyte by the analysis of the ICV solution(s) at each wavelength/ mass used.

## ICP-AES SOLUTION STANDARDS FOR THE SUPERFUND CONTRACT LABORATORY PROGRAM

METHOD  
**CLP**  
ILM 05.2

### ILM 05.2 INSTRUMENT CALIBRATION STANDARDS FOR INDUCTIVELY COUPLED PLASMA - ATOMIC EMISSION SPECTROSCOPY (ICP-AES)

#### Calibration Std. I

5 Components  
100 ug/mL in 5% HNO<sub>3</sub>

Be  
Cd  
Mn  
Pb  
Zn

Part # 52149 \$65/100 mL  
Part # 53149 \$130/500 mL

#### Calibration Std. II

5 Components  
100 ug/mL in 5% HNO<sub>3</sub>

Ba  
Co  
Cu  
Fe  
V

Part # 52150 \$65/100 mL  
Part # 53150 \$130/500 mL

#### Calibration Std. III

2 Components  
100 ug/mL in 5% HNO<sub>3</sub>

As  
Se

Part # 52151 \$40/100 mL  
Part # 53151 \$80/500 mL

#### Calibration Std. IV

6 Components  
100 ug/mL in 5% HNO<sub>3</sub>

Al  
Ca  
Cr  
K  
Na  
Ni

Part # 52152 \$65/100 mL  
Part # 53152 \$130/500 mL

#### Calibration Std. V

4 Components  
100 ug/mL in 5% HNO<sub>3</sub>

Ag  
Mg  
Sb  
Tl

Part # 52153 \$65/100 mL  
Part # 53153 \$130/500 mL

METHOD

**CLP**  
ILM 05.2**ICP-AES SOLUTION STANDARDS FOR THE SUPERFUND  
CONTRACT LABORATORY PROGRAM****ILM 05.2 INITIAL CALIBRATION VERIFICATION (ICV) SOLUTIONS &  
CONTINUING CALIBRATION VERIFICATION (CCV) SOLUTIONS FOR ICP-AES****ICP-AES  
Initial Verification Std. 1**

12 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ba	100
Be	40
Cd	50
Co	100
Cu	100
Fe	100
Pb	100
Mn	100
Ni	100
Ag	20
Tl	100
Zn	100

**Part # 52017      \$95/100 mL****Part # 53017      \$190/500 mL****ICP-AES  
Initial Verification Std. 2**

10 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	100
Sb	100
As	100
Ca	1000
Cr	100
Mg	1000
K	1000
Se	100
Na	1000
V	100

**Part # 52018      \$95/100 mL****Part # 53018      \$190/500 mL****ICP-AES  
Continuing Calibration Verification Standard- CCV**

22 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)	Element	Conc.(ug/mL)
Ca	2500	Zn	20
Mg	2500	Mn	15
K	2500	Cr	10
Na	2500	Ag	10
Al	200	Be	10
Ba	200	Cd	10
Fe	100	Pb	10
Co	50	As	10
V	50	Se	10
Ni	40	Tl	10
Cu	25	Sb	60

**Part # 52154      \$95/100 mL****Part # 53154      \$190/500 mL**

## ICP-AES SOLUTION STANDARDS FOR THE SUPERFUND CONTRACT LABORATORY PROGRAM

METHOD  
**CLP**  
ILM 05.2

### ILM 05.2 CONTRACT REQUIRED QUANTITATION LIMITS (CRQL's) FOR ICP-AES

A standard must be run at double the Contract Required Quantitation Limit (CRQL) or at two times the Instrument Detection Limits (IDL), whichever is greater. This standardization must be performed at the beginning and end of each batch of samples; or at least twice in an 8 hour shift. All elements to be analyzed must be verified except Al, Ba, Ca, Fe, Mg, Na, and K. The CRI Mix 1 contains all the required elements on the Target Analyte List (TAL) in their appropriate concentration ratios. The standard should be diluted by a factor of 1000 prior to use in the "two times CRI" run for ICP analysis. CRI Mix 2 is designed for ICPs equipped with ultrasonic nebulizers and offers all the elements of interest at two times the CRQL level.

<b>CRI Mix 1</b>		<b>CRI Mix 2</b>	
15 Components		15 Components	
Matrix 5% HNO <sub>3</sub>		Matrix 5% HNO <sub>3</sub>	
Element	Conc.(ug/mL)	Element	Conc.(ug/mL)
Sb	120	Sb	120
As	120	As	20
Be	10	Be	10
Cd	10	Cd	10
Cr	20	Cr	20
Co	100	Co	100
Cu	50	Cu	50
Pb	120	Pb	6
Mn	30	Mn	30
Ni	80	Ni	80
Se	120	Se	10
Ag	20	Ag	20
Tl	120	Tl	20
V	100	V	100
Zn	40	Zn	40
<b>Part # 52022</b>	<b>\$125/100 mL</b>	<b>Part # 52023</b>	<b>\$125/100 mL</b>
<b>Part # 53022</b>	<b>\$250/500 mL</b>	<b>Part # 53023</b>	<b>\$250/500 mL</b>

METHOD

**CLP**  
ILM 05.2

**ICP-AES SOLUTION STANDARDS FOR THE SUPERFUND  
CONTRACT LABORATORY PROGRAM**

**ILM 05.2 INTERFERENCE CHECKS  
FOR ICP-AES**

Inter-element and background correction factors must be verified at the beginning and end of each analysis run, or at least twice in an 8 hour shift. Two solutions are required for the interference check: the interferent alone (solution "A"), and the combination of interferents and analytes (solution "AB"). "A" is prepared by diluting Interferents A by a factor of 5. Solution "AB" is prepared by producing a combined solution of Interferents A diluted by a factor of 5 and Analytes B diluted by a factor of 100. For example, combine 20 mL of Interferents A with 1.0 mL of analytes B and dilute the mixture to 100 mL with the calibration blank.

**Interferents A**

4 Components

Matrix 5% HNO<sub>3</sub>.

Element	Conc.(ug/mL)
Al	2500
Ca	2500
Fe	1000
Mg	2500

**Part # 52024      \$95/100 mL**  
**Part # 53024      \$190/500 mL**

**Interferents B**

12 Components

Matrix 5% HNO<sub>3</sub>.

Element	Conc.(ug/mL)
Ag	100
Ba	50
Be	50
Cd	100
Co	50
Cr	50
Cu	50
Mn	50
Ni	100
Pb	100
V	50
Zn	100

**Part # 52025      \$95/100 mL**  
**Part # 53025      \$190/500 mL**

**Spike Sample  
Standard 1**

18 Components

Matrix 5% HNO<sub>3</sub>.

Element	Conc.(ug/mL)
Al	200
Sb	50
As	200
Ba	200
Be	5
Cd	5
Cr	20
Co	50
Cu	25
Fe	100
Pb	50
Mn	50
Ni	50
Se	200
Ag	5
Tl	200
V	50
Zn	50

**Part # 52028      \$280/100 mL**  
**Part # 53028      \$560/500 mL**

**ILM 05.2 SPIKE SAMPLE  
FOR ICP-AES**

In Spike Sample Analysis, a spike containing the required elements in their respective required amounts, is added to the sample prior to addition of any reagents, digestion, distillation, etc.

**Spike Sample  
Standard 3 (Soil)**

16 Components

Matrix 5% HNO<sub>3</sub>.

Element	Conc.(ug/mL)
Sb	100
As	400
Ba	400
Be	10
Cd	10
Cr	40
Co	100
Cu	50
Pb	100
Mn	100
Ni	100
Se	400
Ag	10
Tl	400
V	100
Zn	100

**Part # 52031      \$160/100 mL**  
**Part # 53031      \$320/500 mL**

## ICP-MS SOLUTION STANDARDS FOR THE SUPERFUND CONTRACT LABORATORY PROGRAM

METHOD

### CLP ILM 05.2

### ILM 05.2 CALIBRATION, TUNING, & INTERNAL STANDARDS FOR INDUCTIVELY COUPLED PLASMA - MASS SPECTROSCOPY (ICP-MS)

#### Tuning Standard

5 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Be	100
Mg	100
Co	100
In	100
Pb	100

Part # 52156      \$45/100 mL

Part # 53156      \$90/500 mL

#### ICP-MS

#### Calibration Standard

23 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)	Element	Conc.(ug/mL)
Al	100	Pb	100
Sb	100	Mg	100
As	100	Mn	100
Ba	100	Ni	100
Be	100	Sc	100
Bi	100	Se	100
Cd	100	Ag	100
Cr	100	Tb	100
Co	100	Tl	100
Cu	100	V	100
In	100	Y	100
		Zn	100

Part # 52155      \$95/100 mL

Part # 53155      \$190/500 mL

#### INTERNAL STANDARDS

Element	Matrix	Conc. (ug/mL)	Part#	\$/100mL	Part#	\$/500mL
Bismuth	Bi	Bi(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	1000	57083	25	58083      65
Indium	In	In <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	1000	57049	25	58049      65
Holmium	Ho	Ho <sub>3</sub> O <sub>3</sub> /HNO <sub>3</sub>	1000	57067	25	58067      75
Lithium	Li	LiNO <sub>3</sub> /HNO <sub>3</sub>	1000	57003	25	58003      65
Lutetium	Lu	Lu <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	1000	57071	30	58071      75
Rhodium	Rh	RhCl <sub>3</sub> /HCl	1000	57045	125	58045      400
Scandium	Sc	Sc(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	1000	57021	75	58021      150
Terbium	Tb	Tb <sub>4</sub> O <sub>7</sub> /HNO <sub>3</sub>	1000	57065	25	58065      75
Yttrium	Y	Y <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	1000	57039	25	58039      65

Call Toll-Free 800-368-1131

METHOD

**CLP**  
ILM 05.2

**ICP-MS SOLUTION STANDARDS FOR THE SUPERFUND  
CONTRACT LABORATORY PROGRAM**

**ILM 05.2 INITIAL CALIBRATION VERIFICATION (ICV) SOLUTIONS &  
CONTINUING CALIBRATION VERIFICATION (CCV) SOLUTIONS FOR ICP-MS**

ICP-MS ICV Standard		ICP-MS CCV Standard	
23 Components Matrix 5% HNO <sub>3</sub>		23 Components Matrix 5% HNO <sub>3</sub>	
Element	Conc.(ug/mL)	Element	Conc.(ug/mL)
Al	50	Al	10
Sb	50	Sb	10
As	50	As	10
Ba	50	Ba	10
Be	50	Be	10
Bi	50	Bi	10
Cd	50	Cd	10
Cr	50	Cr	10
Co	50	Co	10
Cu	50	Cu	10
In	50	In	10
Pb	50	Pb	10
Mg	50	Mg	10
Mn	50	Mn	10
Ni	50	Ni	10
Sc	50	Sc	10
Se	50	Se	10
Ag	50	Ag	10
Tb	50	Tb	10
Tl	50	Tl	10
V	50	V	10
Y	50	Y	10
Zn	50	Zn	10


Part # 52157      \$95/100 mL

Part # 53157      \$190/500 mL

Part # 52158      \$95/100 mL

Part # 53158      \$190/500 mL

P#52157-ICP-MS ICV Standard analyzed on HP 4500 ICP-MS





## ICP-MS SOLUTION STANDARDS FOR THE SUPERFUND CONTRACT LABORATORY PROGRAM

METHOD  
**CLP**  
ILM 05.2

### ILM 05.2 CONTRACT REQUIRED QUANTITATION LIMITS (CRQL's) & INTERFERENTS SOLUTIONS FOR ICP-MS

#### ICP-MS - CRI Standard

23 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)	Element	Conc.(ug/mL)
Al	0.1	Pb	0.1
Sb	0.1	Mg	0.1
As	0.1	Mn	0.1
Ba	0.1	Ni	0.1
Be	0.1	Sc	0.1
Bi	0.1	Se	0.1
Cd	0.1	Ag	0.1
Cr	0.1	Tb	0.1
Co	0.1	Tl	0.1
Cu	0.1	V	0.1
In	0.1	Y	0.1
		Zn	0.1

**Part # 52159    \$95/100 mL**  
**Part # 53159    \$190/500 mL**

#### ICP-MS / Interferents A

12 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	100
Ca	100
Fe	100
Mg	100
K	100
Na	100
P	100
SO <sub>4</sub>	100
C	200
Cl	1000
Mo	2
Ti	2

**Part # 52160    \$95/100 mL**  
**Part # 53160    \$190/500 mL**

#### ICP-MS / Interferents B

20 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	20
As	20
Ba	20
Bi	20
Cd	20
Cr	20
Co	20
Cu	20
In	20
Pb	20
Mn	20
Ni	20
Sc	20
Se	20
Ag	20
Tb	20
Tl	20
V	20
Y	20
Zn	20

**Part # 52161    \$110/100 mL**  
**Part # 53161    \$220/500 mL**

METHOD  
**CLP**  
 ILM 05.2

**ICP-MS SOLUTION STANDARDS FOR THE SUPERFUND  
 CONTRACT LABORATORY PROGRAM**

**ILM 05.2 SPIKE SOLUTION FOR ICP-MS**

In Spike Sample Analysis, a spike containing the required elements in their respective required amounts, is added to the sample prior to addition of any reagents, digestion, distillation, etc.

**ICP-MS / Spike Sample**

17 Components  
 Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	2000
Sb	100
As	40
Ba	2000
Be	50
Cd	50
Cr	200
Co	500
Cu	250
Pb	20
Mn	500
Ni	500
Se	10
Ag	50
Tl	50
V	500
Zn	500
<b>Part # 52162</b>	<b>\$110/100 mL</b>
<b>Part # 53162</b>	<b>\$220/500 mL</b>

Place your order 24 hours a day, 7 days a week using

**LabCart™**,

Absolute's on-line ordering system.

Go to [www.AbsoluteStandards.com](http://www.AbsoluteStandards.com) &  
 Click on "On-Line Ordering" for  
*speedy, simple, and secure* purchasing!



**USEPA CONTRACT LABORATORY PROGRAM  
STATEMENT OF WORK FOR INORGANIC ANALYSIS**

METHOD  
**CLP**  
ILM 05.2

### Cold Vapor Mercury Analysis

Element	Conc. (ug/mL)	Part#	\$/100mL	Part#	\$/500mL
Inorganic Mercury	10	57880	25	58880	50
Inorganic Mercury	20	52009	25	53009	50
Inorganic Mercury	100	57280	25	58280	50
Inorganic Mercury	200	57380	25	58380	50
Inorganic Mercury	1000	57080	25	58080	65
Inorganic Mercury	10000	57180	75	58180	150
Organic Mercury	100	54006	25	NA	NA
Organic Mercury	1000	54170	25	54171	50
Total Mercury	100	54005	25	NA	NA
Total Mercury	1000	54004	25	54168	50
Total Mercury	10000	54178	100	NA	NA

### Total Cyanide Analysis

Analyte	Conc. (ug/mL)	Part#	\$/100mL	Part#	\$/500mL
Simple Cyanide	1000	59017	25	54012	90
Total Cyanide	1000	54150	25	NA	NA
Free & Complexed Cyanide	1000	54165	25	NA	NA

Simple Cyanide is formulated with Potassium Cyanide (KCN), no distillation is required.

Total Cyanide is formulated with Potassium Ferricyanide (III), distillation is required before analysis can be performed.

Free & Complexed Cyanide is formulated with both materials.

METHOD

**CLP**  
ILM 05.3

**ICP-AES SOLUTION STANDARDS FOR THE SUPERFUND  
CONTRACT LABORATORY PROGRAM  
WATER MATRIX**

**ILM 05.3 INITIAL CALIBRATION VERIFICATION (ICV) SOLUTIONS &  
CONTINUING CALIBRATION VERIFICATION (CCV) SOLUTIONS FOR  
ICP-AES/ WATER MATRIX**

**ICP-AES ICV  
Water/Soil Matrix  
22 Components  
Matrix 5% HNO<sub>3</sub>**

Element	Conc.(ug/mL)
Al	200
Sb	60
As	10
Ba	200
Be	5
Cd	5
Ca	5000
Cr	10
Co	50
Cu	25
Fe	100
Pb	10
Mg	5000
Mn	15
Ni	40
K	5000
Se	35
Ag	10
Na	5000
Tl	25
V	50
Zn	60

**Part # 52163 \$150/100 mL**  
**Part # 53163 \$300/500 mL**

**ICP-AES CCV  
Water/Soil Matrix  
20 Components  
Matrix 5% HNO<sub>3</sub>**

Element	Conc.(ug/mL)
Al	100
Sb	30
As	5
Ba	100
Ca	2500
Cr	5
Co	25
Cu	12.5
Fe	50
Pb	5
Mn	7.5
Mg	2500
Ni	20
K	2500
Se	17.5
Ag	5
Na	2500
Tl	12.5
V	25
Zn	30

**Part # 52164 \$150/100 mL**  
**Part # 53164 \$300/500 mL**

**ICP-AES SOLUTION STANDARDS FOR THE SUPERFUND  
CONTRACT LABORATORY PROGRAM  
WATER MATRIX**

METHOD  
**CLP**  
ILM 05.3

**ILM 05.3 CONTRACT REQUIRED QUANTITATION LIMITS (CRQL's) &  
SPIKE SOLUTIONS FOR  
ICP-AES/ WATER MATRIX**

**ICP-AES CRI Mix**

**Water Matrix**

15 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	60
As	10
Be	5
Cd	5
Cr	10
Co	50
Cu	25
Pb	10
Mn	15
Ni	40
Se	35
Ag	10
Tl	25
V	50
Zn	60

**Part # 59370 \$125/100 mL**

**Part # 53165 \$250/500 mL**

**ICP-AES Spike Sample**

**Water Matrix**


18 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	200
Sb	10
As	4
Ba	200
Be	5
Cd	5
Cr	20
Co	50
Cu	25
Fe	100
Pb	2
Mn	50
Ni	50
Se	5
Ag	5
Tl	5
V	50
Zn	50

**Part # 52166 \$125/100 mL**

**Part # 53166 \$250/500 mL**



P#59370-ICP-AES CRI Mix analyzed on HP 4500 ICP-MS

METHOD  
**CLP**  
ILM 05.3

**ICP-AES SOLUTION STANDARDS FOR THE SUPERFUND  
CONTRACT LABORATORY PROGRAM  
WATER MATRIX**

**ILM 05.3 INTERFERENTS SOLUTION FOR ICP-AES/ WATER MATRIX**

**ICP-AES  
Interferents A  
Water Matrix**

4 Components

Matrix 5% HNO<sub>3</sub>.

Element	Conc.(ug/mL)
Al	2500
Ca	2500
Fe	1000
Mg	2500

**Part # 52024 \$95/100 mL**

**Part # 53024 \$190/500 mL**

**ICP-AES  
Interferents B  
Water Matrix**

16 Components

Matrix 5% HNO<sub>3</sub>.


Element	Conc.(ug/mL)
Ag	200
As	100
Ba	500
Be	500
Cd	1000
Co	500
Cr	500
Cu	500
Mn	500
Ni	1000
Pb	50
Sb	600
Se	50
Tl	100
V	500
Zn	1000

**Part # 52167 \$125/100 mL**

**Part # 53167 \$250/500 mL**

Combine  
Interferents Mix A & B  
for ICSAB.  
Dilute Interferents A Mix  
into Interferents B Mix in  
a 1:10 ratio.

P#53024 analyzed on HP 4500/ ICP-MS



**ICP-AES SOLUTION STANDARDS FOR THE SUPERFUND  
CONTRACT LABORATORY PROGRAM  
SOIL MATRIX**

METHOD

**CLP  
ILM 05.3**

**ILM 05.3 INITIAL CALIBRATION VERIFICATION (ICV) SOLUTIONS &  
CONTINUING CALIBRATION VERIFICATION (CCV) SOLUTIONS FOR  
ICP-AES/ SOIL MATRIX**

**ICP-AES ICV  
Water/Soil Matrix  
22 Components  
Matrix 5% HNO<sub>3</sub>**

Element	Conc.(ug/mL)
Al	200
Sb	60
As	10
Ba	200
Be	5
Cd	5
Ca	5000
Cr	10
Co	50
Cu	25
Fe	100
Pb	10
Mg	5000
Mn	15
Ni	40
K	5000
Se	35
Ag	10
Na	5000
Tl	25
V	50
Zn	60

**Part # 52163      \$150/100 mL**  
**Part # 53163      \$300/500 mL**

**ICP-AES CCV  
Water/Soil Matrix  
20 Components  
Matrix 5% HNO<sub>3</sub>**

Element	Conc.(ug/mL)
Al	100
Sb	30
As	5
Ba	100
Ca	2500
Cr	5
Co	25
Cu	12.5
Fe	50
Pb	5
Mn	7.5
Mg	2500
Ni	20
K	2500
Se	17.5
Ag	5
Na	2500
Tl	12.5
V	25
Zn	30

**Part # 52164      \$150/100 mL**  
**Part # 53164      \$300/500 mL**

METHOD  
**CLP**  
ILM 05.3

**ICP-AES SOLUTION STANDARDS FOR THE SUPERFUND  
CONTRACT LABORATORY PROGRAM  
SOIL MATRIX**

**ILM 05.3 CONTRACT REQUIRED QUANTITATION LIMITS (CRQL's) &  
SPIKE SOLUTIONS FOR  
ICP-AES/ SOIL MATRIX**

**ICP-AES CRI  
Soil Matrix**

16 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	600
As	100
Ba	500
Be	50
Cd	50
Cr	100
Co	500
Cu	250
Pb	100
Mn	150
Ni	400
Se	350
Ag	100
Tl	250
V	500
Zn	600

**Part # 52168      \$125/100 mL**  
**Part # 53168      \$250/500 mL**

**ICP-AES Spike  
Soil Matrix**


16 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	200
As	80
Ba	4000
Be	100
Cd	100
Cr	400
Co	1000
Cu	500
Pb	40
Mn	1000
Ni	1000
Se	100
Ag	100
Tl	100
V	1000
Zn	1000

**Part # 52169      \$125/100 mL**  
**Part # 53169      \$250/500 mL**

**See page 30  
for  
Interferents Mixes.**

P#52168 - ICP-AES CRI analyzed on HP 4500/ ICP-MS





**ICP-MS SOLUTION STANDARDS FOR THE SUPERFUND  
CONTRACT LABORATORY PROGRAM  
WATER MATRIX**

METHOD

**CLP  
ILM 05.3**

**ILM 05.3 INITIAL CALIBRATION VERIFICATION (ICV) SOLUTIONS &  
CONTRACT REQUIRED QUANTITATION LIMITS (CRQL'S/CRI) SOLUTIONS FOR  
ICP-MS/ WATER MATRIX**

<b>ICP-MS ICV Water Matrix</b>		<b>ICP-MS / CRQL CHECK STANDARD-CRI Water Matrix</b>	
16 Components Matrix 5% HNO <sub>3</sub>		16 Components Matrix 5% HNO <sub>3</sub>	
Element	Conc.(ug/mL)	Element	Conc.(ug/mL)
Sb	20	Sb	20
As	10	As	10
Ba	100	Ba	100
Be	10	Be	10
Cd	10	Cd	10
Cr	20	Cr	20
Co	10	Co	10
Cu	20	Cu	20
Pb	10	Pb	10
Mn	10	Mn	10
Ni	10	Ni	10
Se	50	Se	50
Ag	10	Ag	10
Tl	10	Tl	10
V	10	V	10
Zn	20	Zn	20
<b>Part # 52170</b>	<b>\$125/100 mL</b>	<b>Part # 52171</b>	<b>\$125/100 mL</b>
<b>Part # 53170</b>	<b>\$250/500 mL</b>	<b>Part # 53171</b>	<b>\$250/500 mL</b>

The CRI and ICV mixes are identical in element and concentration, but are prepared from separate source raw materials as specified by the method.

METHOD  
**CLP**  
ILM 05.3

**ICP-MS SOLUTION STANDARDS FOR THE SUPERFUND  
CONTRACT LABORATORY PROGRAM  
WATER MATRIX**

**ICP-MS / CCV  
Water Matrix**

16 Components  
Matrix 5% HNO<sub>3</sub>.

Element	Conc.(ug/mL)
Sb	10.0
As	5.0
Ba	50
Be	5.0
Cd	5.0
Cr	10.0
Co	5.0
Cu	10.0
Pb	5.0
Mn	5.0
Ni	5.0
Se	25.0
Ag	5.0
Tl	5.0
V	5.0
Zn	10.0

Part # 52172 \$125/100 mL  
Part # 53172 \$250/500 mL

**ICP-MS / Interferents A  
Water Matrix**

12 Components  
Matrix 5% HNO<sub>3</sub>.

Element	Conc.(ug/mL)
Al	1000
Ca	1000
Fe	1000
Mg	1000
K	1000
Na	1000
P as PO <sub>4</sub> <sup>3-</sup>	1000
S as SO <sub>4</sub> <sup>2-</sup>	1000
C	2000
Cl	10000
Mo	20
Ti	20

Part # 52174 \$125/100 mL  
Part # 53174 \$250/500 mL

See page 27 for  
Mercury &  
Cyanide  
Standards

Dilute and Combine  
Interferents Mix A & B  
for Mix ICSAB.  
Interferents A Mix is 10X  
the method specification  
& Interferents B Mix is  
1000X.

**ICP-MS / Spike Sample  
Water Matrix**

16 Components  
Matrix 5% HNO<sub>3</sub>.

Element	Conc.(ug/mL)
Sb	100
As	40
Ba	2000
Be	50
Cd	50
Cr	200
Co	500
Cu	250
Pb	20
Mn	500
Ni	500
Se	10
Ag	50
Tl	50
V	500
Zn	500

Part # 52173 \$125/100 mL  
Part # 53173 \$250/500 mL

**ICP-MS / Interferents B  
Water Matrix**

15 Components  
Matrix 5% HNO<sub>3</sub>.

Element	Conc.(ug/mL)
Sb	20
As	20
Ba	20
Cd	20
Cr	20
Co	20
Cu	20
Pb	20
Mn	20
Ni	20
Se	20
Ag	20
Tl	20
V	20
Zn	20

Part # 52175 \$125/100 mL  
Part # 53175 \$250/500 mL

**ICP-MS SOLUTION STANDARDS FOR THE SUPERFUND  
CONTRACT LABORATORY PROGRAM  
SOIL MATRIX**

METHOD

**CLP  
ILM 06.X**

ILM 06.X has the same target analyte list as ILM 05.3, however specifications for ICP-MS in soil are given.

**ICP-MS ICV  
Soil Matrix**

16 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	100
As	50
Ba	500
Be	50
Cd	50
Cr	100
Co	50
Cu	100
Pb	50
Mn	50
Ni	50
Se	250
Ag	50
Tl	50
V	50
Zn	100

**Part # 52176    \$125/100 mL**  
**Part # 53176    \$250/500 mL**

The CRI and ICV mixes are identical in element & concentration, but are prepared from separate source raw materials as

**ICP-MS CRI  
Soil Matrix**

16 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	100
As	50
Ba	500
Be	50
Cd	50
Cr	100
Co	50
Cu	100
Pb	50
Mn	50
Ni	50
Se	250
Ag	50
Tl	50
V	50
Zn	100

**Part # 52177    \$125/100 mL**  
**Part # 53177    \$250/500 mL**

**ICP-MS CCV  
Soil Matrix**

16 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	50
As	25
Ba	250
Be	25
Cd	25
Cr	50
Co	25
Cu	50
Pb	25
Mn	25
Ni	25
Se	125
Ag	25
Tl	25
V	25
Zn	50

**Part # 52178    \$125/100 mL**  
**Part # 53178    \$250/500 mL**

Method  
**TCLP  
(1311)**

**TOXICITY CHARACTERISTIC LEACHATE PROCEDURE  
(TCLP)**

For use in accordance with the "Toxicity Characteristic Rule Regulatory Levels" issued in the Federal Register 55, 1.1846 March 1990: Method 1311.

**(TCLP) Standard**

7 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
As	25
Ba	500
Cd	5
Cr	25
Pb	25
Se	5
Ag	25
Hg*	20

\*Hg is a separate solution.

**Part # 52006 \$95/100 mL**  
**Part # 53006 \$190/500 mL**

**(TCLP) GFAA Standard**

3 Components  
Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
As	25
Pb	25
Se	5

**Part # 52008 \$50/100 mL**  
**Part # 53008 \$100/500 mL**

**(TCLP) Standard for ICP**

4 Components  
Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ba	500
Cd	5
Cr	25
Ag	25

**Part # 52007 \$65/100 mL**  
**Part # 53007 \$130/500 mL**

**(TCLP) Cold Vapor Standard**

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Hg	20

**Part # 52009 \$30/100 mL**  
**Part # 53009 \$50/500 mL**

**Search our complete  
Products Database on-line at  
[www.AbsoluteStandards.com](http://www.AbsoluteStandards.com)**

**Search by: Component  
# of Components  
Concentration  
EPA Method**

## SYNTHETIC PRECIPITATION LEACHING PROCEDURE

Method  
**TCLP  
(1312)**

Element	Matrix	Part#	Conc. (mg/mL)	\$/100mL	Part#	\$/500mL	
Cadmium	Cd	Cd(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57048	1.0	25	58048	65
Cadmium	Cd	Cd(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57148	10.0	75	58148	150
Lead	Pb	Pb(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57082	1.0	25	58082	65
Lead	Pb	Pb(NO <sub>3</sub> ) <sub>2</sub> /HNO <sub>3</sub>	57182	10.0	75	58182	150

### TCLP 1312 METALS

2 Components

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Cd	1000
Pb	5000

**Part # 52148      \$40/100 mL**

**Part # 53148      \$80/500 mL**



**6010A**  
 REV-1,7/92  
**6010B**  
 REV-2,12/96

**MULTI-ELEMENT SOLUTION STANDARDS FOR  
 SW-846 METHODS - ICP-AES**

**MULTI-ELEMENT CALIBRATION STANDARDS FOR 6010A & 6010B**

**Calibration Std. 1**

6 Components  
 Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Be	50
Cd	150
Pb	500
Mn	100
Se	200
Zn	150

**Part # 52043 \$80/100 mL**  
**Part # 53043 \$160/500 mL**

**Calibration Std. 2**

5 Components  
 Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ba	100
Co	100
Cu	100
Fe	1000
V	100

**Part # 52044 \$80/100 mL**  
**Part # 53044 \$160/500 mL**

**Calibration Std. 3**

2 Components  
 Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
As	500
Mo	100

**Part # 52045 \$60/100 mL**  
**Part # 53045 \$120/500 mL**

**Calibration Std. 4**

8 Components  
 Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	200
Ca	1000
Cr	20
Li	100
Ni	20
K	400
Na	200
Sr	10

**Part # 52046 \$90/100 mL**  
**Part # 53046 \$180/500 mL**

**Calibration Std. 5**

4 Components  
 Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	200
Mg	1000
Ag	50
Tl	200

**Part # 52047 \$80/100 mL**  
**Part # 53047 \$160/500 mL**

**Calibration Std. 6**

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
P	200

**Part # 57315 \$30/100 mL**  
**Part # 58315 \$60/500 mL**

**Calibration Standards Set 1-6**

**Part # 52042 \$310/100 mL**  
**Part # 53042 \$620/500 mL**

## MULTI-ELEMENT SOLUTION STANDARDS FOR SW-846 METHODS - ICP-AES

**6010A**  
REV-1,7/92  
**6010B**  
REV-2,12/96

### INTERFERENCE STANDARDS FOR 6010A & 6010B

Interference Standards are used to set or confirm that the appropriate background correction intervals have been set for sequential ICP spectrometers & that the proper inter-element correction factors are set for simultaneous ICP spectrometer systems.

#### Interference Standards Set

Includes one each of  
Interference Standards 1,2,3,4,5  
**Part # 52038      \$320/100 mL**  
**Part # 53038      \$640/ 500mL**

#### Interference Std. 2

Matrix 2% HNO<sub>3</sub> tr Tartaric acid

Element	Conc.(ug/mL)
Sb	1000
<b>Part # 57051</b>	<b>\$30/100 mL</b>
<b>Part # 58051</b>	<b>\$70/500 mL</b>

#### Interference Std. 1

17 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
As	100
Ba	30
Be	10
Cd	30
Cr	30
Co	30
Cu	30
Pb	100
Mn	20
Hg	5
Ni	30
K	3000
Se	50
Ag	30
Tl	100
V	30
Zn	30
<b>Part # 52039</b>	<b>\$230/100 mL</b>
<b>Part # 53039</b>	<b>\$460/500 mL</b>

#### Interference Std. 3

5 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Li	300
Mo	300
P	1000
Sr	200
Ti	1000
<b>Part # 52050</b>	<b>\$80/100 mL</b>
<b>Part # 53050</b>	<b>\$160/500 mL</b>

#### Interference Std. 4

4 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	400
Ca	2000
Fe	2000
Mg	1000
<b>Part # 52041</b>	<b>\$50/100 mL</b>
<b>Part # 53041</b>	<b>\$100/500 mL</b>

#### Interference Std. 5

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Na	5000
<b>Part # 57911</b>	<b>\$50/100 mL</b>
<b>Part # 58911</b>	<b>\$100/500 mL</b>

**6020/  
6020A**

**MULTI-ELEMENT SOLUTION STANDARDS FOR  
SW-846 METHODS - ICP-MS**

**MIXED CALIBRATION STANDARDS FOR 6020 & 6020A**

Method 6020/ 6020A is applicable to the determination of metals in low concentrations (sub-ppb) in water and waste extracts using ICP-MS.

**EPA Method 6020/ 6020A**

**MIX A**

12 Components	
Matrix 5% HNO <sub>3</sub>	
Element	Conc.(ug/mL)
C	1000
Cl	10000
Al	500
Ca	1500
Fe	1250
Mg	500
Na	1250
P	500
K	500
S	500
Mo	10
Ti	10

**Part # 59219 \$175/100 mL**  
**Part # 59249 \$525/500 mL**

**EPA Method 6020/ 6020A**

**MIX B**

11 Components	
Matrix 5% HNO <sub>3</sub>	
Element	Conc.(ug/mL)
As	1.0
Cd	1.0
Cr	2.0
Co	2.0
Cu	2.0
Mn	2.0
Ni	2.0
Se	1.0
Ag	0.5
V	2.0
Zn	1.0

**Part # 52331 \$125/100 mL**  
**Part # 53331 \$250/500 mL**

**ICP-MS InternalStandards**

Element	Matrix	Part#	(ug/mL)	\$/100mL	Part#	\$/500mL
Bismuth Bi	Bi(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57083	1000	25	58083	65
Indium In	In <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57049	1000	25	58049	65
Holmium Ho	Ho <sub>3</sub> O <sub>3</sub> /HNO <sub>3</sub>	57067	1000	25	58067	75
Lithium Li <sup>6+</sup>	Li <sup>6+</sup> NO <sub>3</sub> /HNO <sub>3</sub>	59021	1000	200	59097	1000
Rhodium Rh	RhCl <sub>3</sub> /HCl	57045	1000	120	58045	400
Scandium Sc	Sc(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	57021	1000	75	58021	150
Terbium Tb	Tb <sub>4</sub> O <sub>7</sub> /HNO <sub>3</sub>	57065	1000	25	58065	75
Yttrium Y	Y <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	57039	1000	25	58039	65



## MULTI-ELEMENT SOLUTION STANDARDS FOR SW-846 METHODS- ICP-AES

METHOD  
**200.7**  
Rev 5.0

### MIXED CALIBRATION STANDARDS FOR 200.7

EPA Method 200.7 covers the Determination of Metals and Trace Elements in Water and Wastes by ICP Atomic Emission Spectroscopy.

#### Mixed Calibration Std. 1

10 Components

Matrix 5% HNO<sub>3</sub> / tr.HF  
Element    Conc.(ug/mL)

Ag	50
As	1000
B	100
Ba	100
Ca	1000
Cd	200
Cu	200
Mn	200
Sb	500
Se	500

**Part # 52335      \$80/100 mL**  
**Part # 53335      \$160/500 mL**

#### Mixed Calibration Std. 2

6 Components

Matrix 5% HNO<sub>3</sub>  
Element    Conc.(ug/mL)

K	2000
Li	500
Mo	1000
Na	1000
Sr	100
Ti	1000

**Part # 52336      \$75/100 mL**  
**Part # 53336      \$150/500 mL**

#### Mixed Calibration Std. 3

4 Components

Matrix 5% HNO<sub>3</sub>  
Element    Conc.(ug/mL)

Ce	200
Co	200
V	200
P	1000

**Part # 52337      \$75/100 mL**  
**Part # 53337      \$150/500 mL**

#### Mixed Calibration Std. 4

6 Components

Matrix 5% HNO<sub>3</sub>  
Element    Conc.(ug/mL)

Al	1000
Cr	500
Hg	200
Si	1000
Sn	400
Zn	500

**Part # 52056      \$80/ 100 mL**  
**Part # 53056      \$160/ 500 mL**

#### Mixed Calibration Std. 5

6 Components

Matrix 5% HNO<sub>3</sub>  
Element    Conc.(ug/mL)

Be	100
Fe	1000
Pb	1000
Mg	1000
Ni	200
Tl	500

**Part # 52057      \$110/100 mL**  
**Part # 53057      \$220/500 mL**

METHOD

**200.7**

Various Revisions

**MULTI-ELEMENT SOLUTION STANDARDS FOR  
SW-846 METHODS - ICP-AES**

The **Plasma Solution** is used to determine the optimum height above the load coil for viewing the plasma in EPA Method 200.7, Section 7.6.

The **Tuning Solution** is used to determine the optimum height above the load coil for viewing the plasma in EPA Method 200.7, Section 7.7.

**Plasma Solution**

4 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
As	100
Pb	100
Se	200
Tl	200

**Part # 52058 \$50/100 mL**  
**Part # 53058 \$100/500 mL**

**Tuning Solution**

2 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Cu	100
Pb	100

**Part # 52059 \$40/100 mL**  
**Part # 53059 \$80/500 mL**

The **Laboratory Performance Check Solution Set** is used to evaluate the performance of the instrument system in EPA Method 200.7, Section 7.8.

**LPC Solution Set**

Matrix 5% HNO<sub>3</sub> , Dilution : 1 to 100

**P#52306 LPC-1**

Element	Conc. (ug/mL).
Ag	50
Al	200
As	200
B	200
Ba	200
Be	200
Ca	200
Cd	200
Co	200
Cr	200

**P#52307 LPC-2**

Element	Conc. (ug/mL).
Cu	200
Fe	200
Hg	200
K	1000
Li	200
Mg	200
Mn	200
Mo	200
Na	200
Ni	200

**P#52308 LPC-3**

Element	Conc. (ug/mL).
P	1000
Pb	200
Sb	200
Se	1000
Si	200
Sn*	200
Sr	200
Tl	200
V	200
Zn	200

\* (supplied as a separate solution)

**Part # 52060 \$400/100 mL**  
**Part # 53060 \$800/500 mL**

## MULTI-ELEMENT SOLUTION STANDARDS FOR SW-846 METHODS - ICP-AES

METHOD  
**200.7**  
Various Revisions

The **Spectral Interference Check** Solutions consist of higher-level, concentrated, selected method analytes and are used to evaluate the laboratory's procedure for correcting known inter-element spectral interferences.

**SIC 1** is used to evaluate the Mo inter-element spectral correction factors on the analytes Al, Sb, Se, Sn, and V.

### SIC 1

Matrix H<sub>2</sub>O.  
Dilution : 1 to 10

Element	Conc.(ug/mL)
Mo	500

**Part # 57642      \$30/100 mL**  
**Part # 58642      \$60/500 mL**

**SIC 2** is used to evaluate the Co, Cr, Mn, V, Cu, inter-element spectral correction factors on the analytes Pb, Sb, Mo, As, V, and Zn.

### SIC 2

5 Components  
Matrix 2% HNO<sub>3</sub>.  
Dilution : 1 to 10

Element	Conc. (ug/mL)	Corresponding Interference
Co	100	Pb
Cr	200	Sb
Mn	200	Mo
V	200	As, Be
Cu	400	Zn

**Part # 52061      \$50/100 mL**  
**Part # 53061      \$100/500 mL**

**SIC 3** is used to evaluate the Ni, Al, Fe, inter-element spectral correction factors on the analytes Sb, Zn, As, Ag, Mn, V, Cr.

### SIC 3

3 Components  
Matrix 2% HNO<sub>3</sub>.  
Dilution : 1 to 10

Element	Conc. (ug/mL)	Corresponding Interference
Ni	200	Sb, Zn
Al	300	As
Fe	1500	Ag, Cr, Mn, V

**Part # 52062      \$40/100 mL**  
**Part # 53062      \$80/500 mL**

METHOD

**200.7**

Various Revisions

**MULTI-ELEMENT SOLUTION STANDARDS FOR  
SW-846 METHODS - ICP-AES**

**LABORATORY FORTIFYING SOLUTION (LFS) FOR 200.7**

The Laboratory Fortifying Solution is used in the preparation of the laboratory-fortified blank and the laboratory-fortified sample matrix solutions. The Laboratory-Fortified Blank (LFB) is an aliquot of reagent grade water to which known quantities of the method analytes are added in the laboratory. The LFB is analyzed exactly like a sample to determine whether method performance is within acceptable control limits. The Laboratory-Fortified Sample Matrix (LFM) is an aliquot of an environmental sample to which known quantities of the method analytes are added in the laboratory. The LFM is analyzed exactly like a sample to determine whether the sample matrix contributes bias to the analytical result.


Note:

- The analytes Ca, Mg, and Na are not included in the Fortifying Solution due to their widely varying concentrations in environmental samples
- The analytes B & Si should be disregarded if the samples are processed and diluted in borosilicate labware due to contamination from borosilicate glass.

<b>LFS Solution Set</b>					
Matrix 5% HNO <sub>3</sub> • Dilution : 1 to 100					
<b>P# 52303 LFS-1</b>		<b>P# 52304 LFS-2</b>		<b>P# 52305 LFS-3</b>	
Element	Conc.(ug/mL)	Element	Conc. (ug/mL)	Element	Conc.(ug/mL)
Ag	25	Cu	250	Sb	250
Al	250	Fe	250	Se	250
As	250	Hg	50	Si	250
B	250	K	250	Sn*	100
Ba	250	Li	250	Sr	250
Be	50	Mn	100	Tl	250
Cd	100	Mo	250	V	100
Co	100	Ni	500	Zn	250
Cr	250	Pb	250		

\* (supplied as a separate solution)

**Part # 52063      \$180/100 mL**  
**Part # 53063      \$360/500 mL**



P#52304 - LFS Solution #2 analyzed on HP 4500/ ICP-MS

## MULTI-ELEMENT SOLUTION STANDARDS FOR SW-846 METHODS - ICP-AES

METHOD  
**200.7**  
Various Revisions

### SPIKING STANDARDS FOR 200.7

#### Spiking Standard 1

3 Components

Matrix H<sub>2</sub>O

Element	Conc.(ug/mL)
B	500
Mo	500
Si	2000

Part # 52064 \$80/100 mL

Part # 53064 \$160/500 mL

#### Spiking Standard 4

Matrix 2% HNO<sub>3</sub> tr Tartaric Acid

Element	Conc.(ug/mL)
Sb	500

Part # 57651 \$30/100 mL

Part # 58651 \$60/500 mL

#### Spiking Standard 2

4 Components

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ca	1000
Mg	2000
K	10,000
Na	3000

Part # 52065 \$90/100 mL

Part # 53065 \$180/500 mL

#### Spiking Standard 5

5 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
As	2000
Cd	50
Pb	500
Se	2000
Tl	2000

Part # 52067 \$100/100 mL

Part # 53067 \$200/500 mL

#### Spiking Standard 3

12 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	2000
Ba	2000
Be	50
Cr	200
Co	500
Cu	250
Fe	1000
Mn	500
Ni	500
Ag	50
V	500
Zn	500

Part # 52066 \$180/100 mL

Part # 53066 \$360/500 mL

#### Spiking Standards Set

Includes one each of  
Spiking Standards 1,2,3,4,5

Part # 52068 \$400/100 mL

Part # 53068 \$800/ 500mL

#### Mercury Standard

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Hg	200

Part # 57380 \$30/100 mL

Part # 58380 \$60/500 mL

## METHOD

**200.7**

Various Revisions

**MULTI-ELEMENT SOLUTION STANDARDS FOR  
SW-846 METHODS - ICP-AES****Spiking Standard 1**

4 Components

Matrix H<sub>2</sub>O

Element	Conc.(ug/mL)
B	400
Mo	200
Si	2000
P	400

**Part # 52069 \$90/100 mL****Part # 53069 \$180/500 mL****Spiking Standard 2**

4 Components

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ca	10,000
Mg	10,000
K	10,000
Na	10,000

**Part # 52070 \$120/100 mL****Part # 53070 \$240/500 mL****Spiking Standard 4**Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	200

**Part # 57351 \$30/100 mL****Part # 58351 \$60/500 mL****Spiking Standard 5**

5 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
As	200
Cd	100
Pb	200
Se	400
Tl	400

**Part # 52071 \$100/100 mL****Part # 53071 \$200/500 mL****Spiking Standard 3**

15 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	200
Ba	200
Be	100
Cr	100
Co	200
Cu	200
Fe	400
Li	200
Mn	100
Ni	200
Ag	200
Sr	400
Sn	200
V	200
Zn	200

**Part # 52072 \$180/100 mL****Part # 53072 \$360/500 mL****Mercury Standard**Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Hg	200

**Part # 57380 \$30/100 mL****Part # 58380 \$60/500 mL****Spiking Standards Set**Includes one each of  
Spiking Standards 1,2,3,4,5 &  
Hg @ 200 ug/mL**Part # 52073 \$400/100 mL****Part # 53073 \$800/ 500mL**

## MULTI-ELEMENT SOLUTION STANDARDS FOR SW-846-200.7 & SDWA 1310, NPDWR

METHOD  
**200.7**  
Various Revisions

### DRINKING WATER STANDARDS

These standards are for use in drinking water compliance monitoring and for the analysis of ground and surface water requiring determination at drinking water contaminant levels. Refer to the US EPA SW-846, Method 1310 and US National Primary Drinking Water Regulations 40 CFR Part 141. The Second Drinking Water Metals Standard includes all metals required for US Appendix 200.7 for analysis by ICP Spectroscopy.

#### Primary Drinking Water Metals


8 Components

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
As	10
Ba	100
Cd	5
Cr	10
Pb	10
Se	5
Ag	10
Hg	10

Part # 52075 \$80/100 mL

Part # 53075 \$160/500 mL



P#53075 - Primary Drinking Water Metals verified on HP 4500 ICP-MS

#### Secondary Drinking Water Metals

4 Components

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Cu	100
Fe	30
Mn	5
Zn	500

Part # 52076 \$65/100 mL

Part # 53076 \$130/500 mL


#### Drinking Water Standards Set

Includes one each of  
Primary and Secondary Standards

Part # 52074 \$140/100 mL

Part # 53074 \$280/500 mL

Our quality products are inspected by analytical chemists who have experience using a wide variety of instrumentation. Absolute maintains diverse technologies to ensure your calibration standards are accurate, reliable, & traceable. We use a ICP-MS, HPLC, and many wet methods.





## MULTI-ELEMENT SOLUTION STANDARDS FOR SW-846 METHODS 200.7 & 6010

METHODS  
**200.7**  
**6010**

### INSTRUMENT WAVELENGTH STANDARDS

For calibration and verification of wavelength accuracy and stability in sequential and simultaneous ICP Spectrometers. Also for use in US EPA SW-846, Method 6010 section 5.6 and Method 200.7 section 7.6.1.


#### Instrument Wavelength Standard

15 Components  
Matrix 2% HNO<sub>3</sub>

Al	100	Mg	100
As	100	Mn	100
Cd	100	Ni	100
Cr	100	K	100
Co	100	Na	100
Cu	100	Zn	100
Fe	100	Y	600
Pb	100		

**Part # 52081**      **\$110/100 mL**

**Part # 53081**      **\$220/500 mL**



P#52081 - 200.7 Instrument Wavelength Standard analyzed on HP 4500/ ICP-MS

#### Wavelength Calibration Solution 3

11 Components  
Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
As	20
La	20
Li	20
Cr	20
Mn	20
Mo	20
Ni	20
P	100
K	100
Sc	20
Na	20

**Part # 52079**      **\$90/100 mL**

**Part # 53079**      **\$180/500 mL**

#### Wavelength Calibration Solution 3B

11 Components  
Matrix 2% HCl.

Element	Conc.(ug/mL)
As	20
La	20
Li	20
Mn	20
Mo	20
Ni	20
P	100
K	100
Sc	20
Na	20
S	100

**Part # 59030**      **\$75/100 mL**

METHODS  
**200.7**  
**6010**

**MULTI-ELEMENT SOLUTION STANDARDS FOR  
SW-846 METHODS 200.7 & 6010**

**QUALITY CONTROL STANDARDS**

Quality control standards are used to check standard curves, perform inter-element correction and other spectral interference checks. Also used in US EPA Method 200.7 and 600/482-055 “ Technical Additions to Methods for Chemical Analysis of Water and Wastes.” Available both individually and as a complete set.

**QC Standard 1**

19 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	100
As	100
Be	100
Ca	100
Cd	100
Cr	100
Co	100
Cu	100
Fe	100
Pb	100
Mg	100
Mn	100
Mo	100
Ni	100
Se	100
Tl	100
Ti	100
V	100
Zn	100

**Part # 52084 \$175/100 mL**  
**Part # 53084 \$350/500 mL**

**QC Standard 2**


21 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	100
As	100
Be	100
Ca	100
Cd	100
Cr	100
Co	100
Cu	100
Fe	100
Pb	100
Li	100
Mg	100
Mn	100
Mo	100
Ni	100
Se	100
Sr	100
Tl	100
Ti	100
V	100
Zn	100

**Part # 52085 \$190/100 mL**  
**Part # 53085 \$380/500 mL**

P#52084 - 200.7 QC Standard 1 analyzed on HP 4500/ ICP-MS



## MULTI-ELEMENT SOLUTION STANDARDS FOR SW-846 METHODS 200.7 & 6010

METHODS  
**200.7**  
**6010**

### QUALITY CONTROL STANDARDS CONT'D

#### QC Standard 3

15 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	100
Ba	100
Ca	100
Cd	100
Cr	100
Co	100
Cu	100
Fe	100
Pb	100
Mg	100
Mn	100
Ni	100
Na	100
Ti	100
Zn	100

**Part # 52086**    \$170/100 mL  
**Part # 53086**    \$340/500 mL

#### QC Standard 5

7 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	100
Ba	100
B	100
K	1000
Si	500
Ag	50
Na	100

**Part # 52088**    \$80/100 mL  
**Part # 53088**    \$160/500 mL

#### QC Standard 4

7 Components

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	100
Ba	100
B	100
K	1000
Si	50
Ag	100
Na	100

**Part # 52087**    \$80/100 mL  
**Part # 53087**    \$160/500 mL

#### Calibration Standards Set

Includes one each of Quality Control Standard 1, 3, 5.


**Part # 52089**    \$310/100 mL  
**Part # 53089**    \$620/500 mL

#### Calibration Standards Set

Includes one each of Quality Control Standard 2, 3, 4.

**Part # 52090**    \$310/100 mL  
**Part # 53090**    \$620/500 mL

P#52087 - 200.7 QC Standard 4 analyzed on HP 4500/ ICP-MS



METHOD

**200.8****MULTI-ELEMENT SOLUTION STANDARDS FOR  
METHOD 200.8****ICP-MS MULTI-ELEMENT TUNING STANDARDS REV 5.5**

These multi-element standards are designed to assist the operator in calibration of the instrument or verification of the mass range.

**ICP-MS Tuning Standard 1**

5 Components	
Matrix 2% HNO <sub>3</sub>	
Element	Conc.(ug/mL)
Be	10
Co	10
In	10
Pb	10
Mg	10

**Part # 52122 \$50/100 mL**  
**Part # 53122 \$200/500 mL**

**ICP-MS Tuning Standard 2**

4 Components	
Matrix 2% HNO <sub>3</sub>	
Element	Conc.(ug/mL)
Li	10
Y	10
Ce	10
Tl	10

**Part # 52123 \$50/100 mL**  
**Part # 53123 \$200/500 mL**

**ICP-MS INTERNAL STANDARDS**

These internal standard stock solutions are used for: addition to blanks; calibration standards; as samples for internal standardization; and for use in US EPA SW-846 Method 200.8. Dilute 10 mL to 100 mL before using.

Element	Matrix	(ug/mL)	Part#	\$/100mL	Part#	\$/500mL
Bismuth Bi	Bi(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	100	57283	25	58283	50
Indium In	In <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	100	57249	25	58249	50
Holmium Ho	Ho <sub>3</sub> O <sub>3</sub> /HNO <sub>3</sub>	1000	57067	25	58067	75
Lithium Li <sup>6+</sup>	Li <sup>6+</sup> NO <sub>3</sub> /HNO <sub>3</sub>	100	59036	100	59037	300
Rhodium Rh	RhCl <sub>3</sub> /HCl	100	57245	100	58245	300
Scandium Sc	Sc(NO <sub>3</sub> ) <sub>3</sub> /HNO <sub>3</sub>	100	57221	25	58221	50
Terbium Tb	Tb <sub>4</sub> O <sub>7</sub> /HNO <sub>3</sub>	100	57265	25	58265	50
Yttrium Y	Y <sub>2</sub> O <sub>3</sub> /HNO <sub>3</sub>	100	57239	25	58239	50

**Method 200.8 Rev 5.5  
ICP-MS Internal Standard**

5 Components	
Matrix 2% HNO <sub>3</sub>	
Element	Conc.(ug/mL)
Sc	10
Y	10
In	10
Tb	10
Bi	10

**Part # 52332 \$50/100 mL**  
**Part # 53332 \$200/500 mL**

**Multi-Element  
Internal Standard**

7 Components	
Matrix 2% HNO <sub>3</sub>	
Element	Conc.(ug/mL)
Bi	10
Ho	10
In	10
Li	10
Sc	10
Tb	10
Y	10

**Part # 52319 \$65/100 mL**  
**Part # 53319 \$200/500 mL**

<b>MULTI-ELEMENT SOLUTION STANDARDS FOR METHOD 200.8</b>	<b>METHOD 200.8</b>
--------------------------------------------------------------	-------------------------

**ICP-MS CALIBRATION STANDARDS REV 5.5**

These multi-element standards are designed to assist the ICP-MS operator in performing concentration verification

**Method 200.8  
ICP-MS  
Calibration Standard 1**

19 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	10
Sb	10
As	10
Be	10
Cd	10
Cr	10
Co	10
Cu	10
Pb	10
Mn	10
Hg	5
Mo	10
Ni	10
Se	50
Tl	10
Th	10
U	10
V	10
Zn	10

**Part # 52333      \$150/100 mL**  
**Part # 53333      \$300/500 mL**

**Method 200.8  
ICP-MS  
Calibration Standard 3**

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Hg	5

**Part # 59388      \$30/100 mL**


**Method 200.8  
ICP-MS  
Calibration Standard 2**

2 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ba	10
Ag	10

**Part # 52334      \$40/100 mL**  
**Part # 53334      \$80/500 mL**

P#53332 - 200.8 Internal Standard analyzed on HP 4500/ ICP-MS



METHOD

**MULTI-ELEMENT SOLUTION STANDARDS FOR  
METHOD 200.8**

**200.8**

**ICP-MS VERIFICATION STANDARD**

**ICP-MS Verification  
Standard 1**

26 Components  
Matrix 5% HNO<sub>3</sub>

Element	(ug/mL)	Element	(ug/mL)
Al	10	Li	10
As	10	Mg	10
Ba	10	Mn	10
Be	10	Hg	10*Separate solution
Bi	10	Ni	10
Ca	10	K	10
Cd	10	Se	10
Cr	10	Na	10
Co	10	Ag	10
Cu	10	Sr	10
Fe	10	Tl	10
Ga	10	V	10
Pb	10	Zn	10

**Part # 52100 \$180/100 mL**  
**Part # 53100 \$380/500 mL**  
 \*Part# 57880 is included (Hg @ 10 ug/mL)

**ICP-MS Verification  
Standard 2**

6 Components  
Matrix H<sub>2</sub>O

Element	Conc.(ug/mL)
B	10
Mo	10
S	10
Si	10
Ti	10
W	10


**Part # 52101 \$60/100 mL**  
**Part # 53101 \$120/500 mL**

**ICP-MS Verification Set**

Includes one each of ICP-MS Veri-  
fication 1 and 2.

**Part # 52102 \$220/100 mL**  
**Part # 52302 \$440/500 mL**

P#52100- 200.8 Verification Standard 1 analyzed on HP 4500/ ICP-MS



## MULTI-ELEMENT SOLUTION STANDARDS FOR METHOD 200.8

METHOD

**200.8**

### ICP-MS INTERFERENCE CHECK SOLUTIONS

#### ICP-MS Interference Check Solution A

12 Components	
Matrix in 2%-HNO <sub>3</sub> / tr. HF	
Element	Conc.(ug/mL)
Al	500
Ca	500
Fe	500
Mg	500
Na	500
P	500
K	500
S	500
C	1000
Cl	3600
Mo	10
Ti	10

**Part # 52114**     **\$100/100 mL**  
**Part # 53114**     **\$450/500 mL**

#### ICP-MS Interference Check Solution B

11 Components	
Matrix in 2%-HNO <sub>3</sub> / tr. HF	
Element	Conc.(ug/mL)
As	0.100
Cd	0.050
Cr	0.100
Co	0.200
Cu	0.100
Ni	0.200
Mn	0.100
Se	0.100
Ag	0.100
V	0.200
Zn	0.100

**Part # 52115**     **\$100/100 mL**  
**Part # 53115**     **\$450/500 mL**

### ICP-MS MEMORY CHECK SOLUTIONS

Combine mixes A and B in equal ratios.

This test will determine any memory problems that will affect the quality of the data.

#### ICP-MS Memory Check Solution A

14 Components	
Matrix in 2%-HNO <sub>3</sub> / tr. HF	
Element	Conc.(ug/mL)
Al	1000
Mg	1000
Na	1000
C	2000
Cl	7200
S	1000
Ti	20
Sb	20
Be	20
Cr	20
Mn	20
Ni	20
V	20
Zn	20

**Part # 52116**     **\$100/100 mL**  
**Part # 53116**     **\$450/500 mL**

#### ICP-MS Memory Check Solution B

14 Components	
Matrix in 2%-HNO <sub>3</sub>	
Element	Conc.(ug/mL)
Ca	1000
Fe	1000
K	1000
P	1000
Mo	20
As	20
Ba	20
Cd	20
Co	20
Cu	20
Pb	20
Se	20
Ag	20
Tl	20

**Part # 52117**     **\$100/100 mL**  
**Part # 53117**     **\$450/500 mL**

## ICP-MS

MULTI-ELEMENT SOLUTION STANDARDS  
FOR ICP-MS**Multi-Element Solution 1**

29 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ag	10
Al	10
As	10
Ba	10
Be	10
Bi	10
Ca	10
Cd	10
Co	10
Cr	10
Cs	10
Cu	10
Fe	10
Ga	10
In	10
K	10
Li	10
Mg	10
Mn	10
Na	10
Ni	10
Pb	10
Rb	10
Se	10
Sr	10
Tl	10
U	10
V	10
Zn	10

**Part # 52133**    **\$180/100 mL**  
**Part # 53133**    **\$360/500 mL**

**Multi-Element Solution 2**

17 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ce	10
Dy	10
Er	10
Eu	10
Gd	10
Ho	10
La	10
Lu	10
Nd	10
Pr	10
Sc	10
Sm	10
Tb	10
Th	10
Tm	10
Y	10
Yb	10

**Part # 52134**    **\$200/100 mL**  
**Part # 53134**    **\$400/500 mL**

**Multi-Element Solution 3**

10 Components

Matrix 10% HCL / 1% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Au	10
Hf	10
Ir	10
Pd	10
Pt	10
Rh	10
Ru	10
Sb	10
Sn	10
Te	10

**Part # 52135**    **\$250/100 mL**  
**Part # 53135**    **\$500/500 mL**

**Multi-Element Solution 4**

12 Components

Matrix H<sub>2</sub>O tr. HF

Element	Conc.(ug/mL)
B	10
Ge	10
Mo	10
Nb	10
P	10
Re	10
S	10
Si	10
Ta	10
Ti	10
W	10
Zr	10

**Part # 52136**    **\$150/100 mL**  
**Part # 53136**    **\$300/500 mL**

**Multi-Element Solution 5**Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Hg	10

**Part # 57880**    **\$30/100 mL**  
**Part # 58880**    **\$60/500 mL**

**ICP-MS Solution Kit**Includes one each of  
Multi-Element 1, 2, 3, 4, & 5.

**Part # 52137**    **Part # 53137**  
**5X100 mL**      **5X500 mL**  
**\$650**            **\$1300**



## DETERMINATION OF TRACE ELEMENTS BY STABILIZED TEMPERATURE GRAPHITE FURNACE AA

METHOD  
**200.9**

### Method 200.9 Complete Single Element Kit (16 x 100ml bottles @ 1000ug/mL Each)

Element	Matrix
Aluminum	Al $\text{Al}(\text{NO}_3)_3/\text{HNO}_3$
Antimony	Sb $\text{Sb}_2\text{O}_3/\text{HNO}_3$ tr. Tartaric acid
Arsenic	As $\text{As}_2\text{O}_3/\text{HNO}_3$
Beryllium	Be $\text{Be}_4\text{O}(\text{C}_2\text{H}_3\text{O}_2)_6/\text{HNO}_3$
Cadmium	Cd $\text{Cd}(\text{NO}_3)_2/\text{HNO}_3$
Chromium	Cr $\text{Cr}(\text{NO}_3)_3/\text{HNO}_3$
Cobalt	Co $\text{Co}(\text{NO}_3)_2/\text{HNO}_3$
Copper	Cu $\text{Cu}(\text{NO}_3)_2/\text{HNO}_3$
Iron	Fe $\text{Fe}(\text{NO}_3)_3/\text{HNO}_3$
Lead	Pb $\text{Pb}(\text{NO}_3)_2/\text{HNO}_3$
Manganese	Mn $\text{Mn}(\text{NO}_3)_2/\text{HNO}_3$
Nickel	Ni $\text{Ni}(\text{NO}_3)_2/\text{HNO}_3$
Selenium	Se $\text{SeO}_2/\text{HNO}_3$
Silver	Ag $\text{AgNO}_3/\text{HNO}_3$
Thallium	Tl $\text{TlNO}_3/\text{HNO}_3$
Tin	Sn $(\text{NH}_4)_2\text{SnF}_6/\text{HNO}_3/\text{HCl}$

**Part # 52252 \$280/ 16 x 100 mL**

### Palladium and Magnesium Nitrate Modifier

Starting Material Pd &  $\text{Mg}(\text{NO}_3)_2$

Matrix 0.3%(3 mg/mL) Pd & 0.2%(2 mg/mL) Mg in 5%  $\text{HNO}_3$

**Part # 52311 \$250/100mL**

**Part # 53311 \$450/500mL**

#### PT Study Schedule:

**January 15- March 1**

**April 15- May 30**

**\*July 15- August 29 (DMR-QA)**

**October 15- November 29**

**Offering our complete line of  
water & soil**

**PT's EVERY study!**



METHOD

**1620**

## METALS BY INDUCTIVELY COUPLED PLASMA ATOMIC EMISSION SPECTROSCOPY & ATOMIC ABSORPTION SPECTROSCOPY

The method is a consolidation of US EPA Methods 200.7 (ICP for trace elements), 204.2 (Sb), 206.2 (As), 239.2 (Pb), 270.2 (Se), 279.2 (Tl), 245.5 (Hg), 245.1 (Hg), and 245.2(Hg). The method is used for analysis of trace elements by ICP atomic emission spectroscopy and GFAA spectroscopy, for analysis of mercury by CVAA spectroscopy, and as a semi-quantitative ICP screen for specified elements.

**EPA 1620 Standard 1**

5 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Mn	100
Be	100
Cd	100
Pb	100
Zn	100

**Part # 52139     \$60/100 mL****Part # 53139     \$120/500 mL****EPA 1620 Standard 2**

6 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ba	100
Cu	100
Fe	100
V	100
Y	100
Co	100

**Part # 52140     \$60/100 mL****Part # 53140     \$120/500 mL****EPA 1620 Standard 3**

3 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Mo	100
As	100
Se	100

**Part # 52141     \$50/100 mL****Part # 53141     \$100/500 mL****EPA 1620 Standard 4**

5 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ca	100
Na	100
Al	100
Cr	100
Ni	100

**Part # 52142     \$60/100 mL****Part # 53142     \$120/500 mL****EPA 1620 Standard 5**

6 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	100
B	100
Mg	100
Ag	100
Tl	100
Ti	100

**Part # 52143     \$60/100 mL****Part # 53143     \$120/500 mL****EPA 1620 Standard 6****Tin**Matrix 2% HNO<sub>3</sub> / 6% HCL

Element	Conc.(ug/mL)
Sn	100

**Part # 57250     \$30/100 mL****Part # 58250     \$60/500 mL**

**DETERMINATION OF HEXAVALENT CHROMIUM BY  
ION CHROMATOGRAPHY**

METHOD

**1636**

This method is for the determination of dissolved hexavalent chromium (as CrO) in ambient waters at EPA water quality criteria (WQC) levels using ion chromatography (IC). This method was developed by integrating the analytical procedures in EPA Method 218.6 with the quality control (QC) and sample handling procedures necessary to avoid contamination and ensure the validity of analytical results during sampling and analysis for metals at EPA WQC levels.

**Hexavalent Chromium (Cr<sup>6+</sup>)**

1000 ug/mL in Water

**Part # 54161     \$30/100 mL**  
**Part # 54172     \$50/500 mL**

**Hexavalent Chromium (Cr<sup>6+</sup>)**

10 mg/mL in Water

**Part # 54175     \$75/100 mL**  
**Part # 54176     \$150/500 mL**

## METHODS

**1638****1640**
**DETERMINATION OF TRACE ELEMENTS IN AMBIENT  
WATERS BY INDUCTIVELY COUPLED PLASMA -  
MASS SPECTROMETRY**

Method 1638 & 1640 are for the determination of dissolved elements in ambient waters at EPA water quality criteria (WQC) levels using inductively coupled plasma-mass spectrometry (ICP-MS). In addition, Method 1640 employs on-line chelation preconcentration. These methods may also be used for determination of total recoverable element concentrations in these waters.

**Method 1638 Mix A**

8 Components

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	10
Cd	10
Cu	10
Pb	10
Ni	10
Se	10
Tl	10
Zn	10

**Part # 52144**    \$60/100 mL  
**Part # 53144**    \$120/500 mL

**Method 1640 Mix A**

6 Components

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
As	10
Cd	10
Cu	10
Pb	10
Ni	10
Zn	10

**Part # 52146**    \$50/100 mL  
**Part # 53146**    \$100/500 mL

**Method 1638/1640 Mix B****Silver (Ag)**Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ag	20

**Part # 57847**    \$30/100 mL  
**Part # 58847**    \$60/500 mL

**Internal Standard**

5 Components

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sc	10
Y	10
In	10
Tb	10
Bi	10

**Part # 52145**    \$60/100 mL  
**Part # 53145**    \$120/500 mL

**Method 1638  
Tuning Standard**

5 Components

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Be	10
Co	10
In	10
Pb	10
Mg	10

**Part # 52122**    \$50/100 mL  
**Part # 53122**    \$200/500 mL

**Method 1640  
Tuning Standard**

5 Components

Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ni	10
Y	10
In	10
Tb	10
Pb	10

**Part # 52147**    \$50/100 mL  
**Part # 53147**    \$100/500 mL

**N-HEXANE EXTRACTABLE MATERIAL (HEM;  
OIL & GREASE) & SILICA GEL TREATED N-HEXANE  
EXTRACTABLE MATERIAL BY EXTRACTION & GRAVIMETRY**

METHOD

**1664**

As a party to the Montreal Protocol on Substances that Deplete the Ozone Layer and as required by law under the Clean Air Act Amendments of 1990 (CAAA), the United States is committed to controlling and eventually phasing out the use of chlorofluorocarbons (CFCs). In support of these efforts, Method 1664 was developed by the United States Environmental Protection Agency Office of Science and Technology to replace previously used gravimetric procedures that employed Freon-113, a Class I CFC, as the extraction solvent for the determination of oil and grease and non-polar material.

This method is for determination of n-hexane extractable material (HEM; oil and grease) and n-hexane extractable material that is not adsorbed by silica gel (SGT-HEM; non-polar material) in surface and saline waters and industrial and domestic aqueous wastes. Extractable materials that may be determined are relatively non-volatile hydrocarbons, vegetable oils, animal fats, waxes, soaps, greases, and related materials. The method is based on prior Environmental Protection Agency (EPA) methods for determination of “oil and grease” and “total petroleum hydrocarbons”

**EPA Method 1664****Oil & Grease**

8000 (ug/mL) in Acetone

- (1) n-Hexadecane
- (2) Stearic acid

**Part # 91958      \$65/100mL****Also Available****EPA Method 413****Oil & Grease**

1000 ug/mL in n-Propanol/Glycerol.

- (1) Cooking Oil
- (2) Paraffin

**Part # 54135      \$25/100 mL**

METHOD  
**600/4-79-020**

**MULTI-ELEMENT SOLUTION STANDARDS FOR  
GROUNDWATER AND WASTEWATER**

**GROUNDWATER AND WASTEWATER STANDARDS**

These may be used either as standards or as a means of checking an individual analyst's accuracy and precision. Refer to the US EPA Methods Manual, 600/4-79-020, "Methods for Chemical Analysis of Water and Wastes" for the appropriate Trace Metals I, II, III procedures.

**Trace Metals 1**

15 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	500
As	100
Be	100
Cd	25
Cr	100
Co	100
Cu	100
Fe	100
Pb	100
Mn	100
Hg	5
Ni	100
Se	25
V	250
Zn	100

**Part # 52092 \$130/100 mL**

**Part # 53092 \$260/500 mL**

**Trace Metals 2**

3 Components

Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Sb	20
Ag	10
Tl	20

**Part # 52093 \$50/100 mL**

**Part # 53093 \$100/500 mL**

**Trace Metals 3**

6 Components


Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ba	500
Ca	500
Mg	100
Mo	500
K	100
Na	500

**Part # 52094 \$75/100 mL**

**Part # 53094 \$150/500 mL**

P#52092- 600/4-79-020 Trace Metals 1 analyzed on HP 4500/ ICP-MS



**Trace Metals Set**

Includes one each of  
Trace Metals 1, 2, and 3.

**Part # 52091 \$240/100 mL**

**Part # 53091 \$480/500 mL**

## MULTI-ELEMENT SOLUTION STANDARDS FOR TOXIC METALS IN FISH

METHOD  
**600/4-79-020**

### ALTERNATE WATER & WASTEWATER STANDARDS

#### Alternate Metals Set

Includes one each of  
Alternate Metals 1 and 2.

**Part # 52095**      **\$120/100 mL**  
**Part # 53095**      **\$240/500 mL**

#### Alternate Metals 2

4 Components  
Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Ca	500
Mg	100
K	100
Na	500

**Part # 52097**      **\$90/100 mL**  
**Part # 53097**      **\$180/500 mL**

#### Alternate Metals 1


11 Components  
Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Al	20
Sb	5
Be	5
Co	10
Cu	10
Fe	20
Mn	10
Ni	10
Tl	5
V	20
Zn	10

**Part # 52096**      **\$60/100 mL**  
**Part # 53096**      **\$120/500 mL**

### TRACE METALS IN FISH

For use in methods for the sampling and analysis of priority pollutants in sediments and fish tissue. Refer to the US EPA Methods Manuals, 600/4-79-020 and 600/4-81-055 for the appropriate procedures.



#### Trace Metals in Fish

9 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
As	100
Cd	5
Cr	20
Cu	50
Pb	10
Hg	100
Ni	2
Se	10
Zn	1000

**Part # 52098**      **\$90/100 mL**  
**Part # 53098**      **\$180/500 mL**

**MISC.**

**MISCELLANEOUS METAL STANDARDS**

**Misc. Spike Solution #1**

11 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
Co	1000
Sr	1000
Al	1000
Ba	1000
Cd	1000
Cr	1000
Cu	1000
Se	1000
V	1000
Be	1000
As	1000

**Part # 59045      \$250/100 mL**

**Spike Solution Set**

Includes one each of Solution #1 & #2.

**Part # 52259      \$400/ 2x100 mL**

**Misc. Spike Solution #2**

10 Components  
Matrix 5% HNO<sub>3</sub>

Element	Conc.(ug/mL)
B	1000
Fe	1000
Pb	1000
Mn	1000
Mo	1000
Ni	1000
Tl	1000
Zn	1000
Ag	200
Sb	1000

**Part # 59046      \$250/100 mL**


**Metals Mix**

24 Components  
Matrix 2% HNO<sub>3</sub>

Element	Conc.(ug/mL)	Element	Conc.(ug/mL)
Fe	1000	Cr	10
K	1000	Cu	10
Ca	1000	Mn	10
Na	1000	Mo	10
Mg	1000	Ni	10
Ag	10	Pb	10
Al	10	Sb	10
As	10	Se	10
Ba	10	Tl	10
Be	10	V	10
Cd	10	Zn	10
Co	10	Ti	10

**Part # 59110      \$195/100 mL**

**Part # 59153      \$500/500 mL**





## DETERMINATION OF INORGANIC ANIONS BY ION CHROMATOGRAPHY

METHOD

**300.0**

### EPA Method 300.0 Anions Single Components

1000 ug/mL in Water.

Calibration Mixes	Part#	100 mL	Part#	500mL
Chloride	54101	\$25	54501	\$90
Fluoride	54102	\$25	54502	\$90
Nitrate as N	54103	\$25	54503	\$50
Nitrite as N	54104	\$25	54504	\$50
Ortho-phosphate as P	54105	\$25	54505	\$90
Sulfate	54106	\$25	54506	\$90
Bromide	54107	\$25	54507	\$90
Bromate	54108	\$25	54508	\$90
Chlorite	54109	\$25	54509	\$90
Chlorate	54110	\$25	54510	\$90

#### Complete set of all 10 solutions

Part # 54000 \$150/100 mL

Part # 54500 \$750/500 mL

#### Sodium Carbonate

Eluent 100 x Concentrate

0.5 M in Water.

Part # 54127 \$65/100 mL

Part # 54527 \$100/500 mL

#### Sulfide

1000 ug/mL in Water.

Part # 54139 \$40/100 mL

#### Sodium Bicarbonate

Eluent 100 x Concentrate

0.5 M in Water.

Part # 54128 \$25/100 mL

Part # 54528 \$100/500 mL

#### Chloride (Cl<sup>-</sup>)

From CaCl<sub>2</sub>

1000 ug/mL in Water.

Part # 54155 \$25/100 mL

#### EPA Method 300.0 Eluent 100 x Concentrate

Varied Molarity in Water. 100 mL

(1) Sodium carbonate 0.18 M

(2) Sodium bicarbonate 0.17 M

Part # 54129 \$25/100 mL

Part # 54529 \$100/500 mL

METHOD

## DETERMINATION OF INORGANIC ANIONS BY ION CHROMATOGRAPHY

**300.0**

A small volume of sample, typically 2 to 3 mL, is introduced into an ion chromatograph. The anions of interest are separated and measured, using a system comprised of a guard column, separator column, suppressor device, and conductivity detector. The main differences between Method A and B are the separator columns, guard columns and eluents. In order to use this method for solids, an extraction procedure must be performed. Stability of the solutions must be closely monitored. Store the materials at 4°C.

**EPA Method 300.0 Anions****Mix A**

1000 ug/mL in Water.

- (1) Bromide
- (2) Chloride
- (3) Fluoride
- (4) Nitrate as N
- (5) Nitrite as N
- (6) Ortho-phosphate as P
- (7) Sulfate

**Part # 52118    \$50/100 mL****Part # 53118    \$200/500 mL****EPA Method 300.0 Anions****Mix A Alternate**

1000 ug/mL in Water.

- (1) Bromide
- (2) Chloride
- (3) Fluoride
- (4) Nitrate as N
- (5) Ortho-phosphate as P
- (6) Sulfate

**Part # 52320    \$50/100 mL****Part # 53320    \$200/500 mL****EPA Method 300.0 Anions****Mix B**

1000 ug/mL in Water.

- (1) Chlorite
- (2) Chlorate
- (3) Bromate

**Part # 52119    \$25/100 mL****Part # 53119    \$100/500 mL****Anions**

100 ug/mL in Water.

- (1) Nitrate (NO<sub>3</sub><sup>-</sup>)-as N
- (2) Nitrite (NO<sub>2</sub><sup>-</sup>)-as N
- (3) Fluoride (F<sup>-</sup>)
- (4) Orthophosphate (PO<sub>4</sub><sup>3-</sup>)-as P
- (5) Nitrate & Nitrite as N

**Part # 52107    \$50/100 mL****Part # 53107    \$150/500 mL****\*Ohio Monthly QC Calibration Check**

Varied ug/mL in Water.

Fluoride

**Part # 59076    \$25/100 mL****Part # 59176    \$60/500 mL**

\*Meets the Ohio concentration requirement of 0.5 -1.5 mg/L.

## DETERMINATION OF INORGANIC ANIONS BY ION CHROMATOGRAPHY

METHOD

**300.0****EPA Method 300.0 Anions Mix**


Varied ug/mL in Water.

(1)	Bromide	20
(2)	Chloride	20
(3)	Fluoride	10
(4)	Nitrate as N	20
(5)	Nitrite as N	20
(6)	Ortho-phosphate as P	30
(7)	Sulfate	30

**Part # 59010 \$50/100 mL****Part # 59095 \$200/500 mL****EPA Method 300.0 Anions**

Varied ug/mL in Water.

(1)	Bromide	200
(2)	Chloride	200
(3)	Fluoride	100
(4)	Nitrate as N	200
(5)	Nitrite as N	200
(6)	Ortho-phosphate as P	300
(7)	Sulfate	300

**Part # 59011 \$100/100 mL****Part # 59096 \$200/500 mL**

METHOD

**300.0**

**DETERMINATION OF INORGANIC ANIONS BY ION CHROMATOGRAPHY**

**INORGANIC DISINFECTION BY-PRODUCTS**

**Inorganic Chlorinated Disinfection By-Products**

1000 ug/mL in Water.

- (1) Chlorate (ClO<sub>3</sub>)
- (2) Chlorite (ClO<sub>2</sub>)

Part # 52105     \$35/100 mL  
 Part # 53105     \$95/500 mL

**Inorganic Brominated Disinfection By-Products**

1000 ug/mL in Water.

- (1) Bromate (BrO<sub>3</sub>)
- (2) Bromide (Br)

Part # 52106     \$35/100 mL  
 Part # 53106     \$95/500 mL


**EPA Method 300.0 Anions Mix B**

1000 ug/mL in Water.

- (1) Chlorite
- (2) Chlorate
- (3) Bromate

Part # 52119     \$25/100 mL  
 Part # 53119     \$100/500 mL

\*DAD1 A, Sig=360,20 Ref=266,10 (102604|52106A1.D - 102604|BLANK3.D)



## DETERMINATION OF INORGANIC CATIONS BY ION CHROMATOGRAPHY

## Cations

### INORGANIC CATIONS ION CHROMATOGRAPHY

#### Single Components

1000 ug/mL - 100 mL Volume

54112	Ammonium, NH <sub>4</sub> <sup>+</sup>	in Water.	\$25
54113	Ammonium as N	in Water.	\$25
57056	Barium, Ba <sup>2+</sup>	in 2% HNO <sub>3</sub> .	\$30
57020	Calcium, Ca <sup>2+</sup>	in 2% HNO <sub>3</sub> .	\$30
57003	Lithium, Li <sup>+</sup>	in 2% HNO <sub>3</sub> .	\$30
57012	Magnesium, Mg <sup>2+</sup>	in 2% HNO <sub>3</sub> .	\$30
57019	Potassium, K <sup>+</sup>	in 2% HNO <sub>3</sub> .	\$30
57011	Sodium, Na <sup>+</sup>	in 2% HNO <sub>3</sub> .	\$30
57038	Strontium, Sr <sup>2+</sup>	in 2% HNO <sub>3</sub> .	\$30

**Complete set of all 9 solutions,  
use Part # 54001 \$150**

### Additional Inorganic Cations Ion Chromatography

#### Single Components

1000 ug/mL - \$25/ 100 mL

54154	Calcium (Ca <sup>2+</sup> )	in Water.	
54157	Magnesium (Mg <sup>2+</sup> )	in Water.	
54158	Potassium (K <sup>+</sup> )	in Water.	
54149	Sodium (Na <sup>+</sup> )	in Water.	

#### Cations Mix #1

4 Components in Water.

(1)	Calcium (Ca <sup>2+</sup> )	1000
(2)	Magnesium, (Mg <sup>2+</sup> )	1000
(3)	Potassium, (K <sup>+</sup> )	1000
(4)	Sodium (Na <sup>+</sup> )	1000

**Part # 59322      \$50/100 mL**

#### Cations Mix #2

6 Components in Water.

(1)	Calcium (Ca <sup>2+</sup> )	100
(2)	Magnesium, (Mg <sup>2+</sup> )	100
(3)	Potassium, (K <sup>+</sup> )	100
(4)	Sodium (Na <sup>+</sup> )	100
(5)	Ammonium (NH <sub>4</sub> <sup>+</sup> )	100
(6)	Lithium (Li <sup>+</sup> )	50

**Part # 59203      \$85/100 mL**

METHOD  
**314-  
332.0**

**DETERMINATION OF PERCHLORATE BY  
ION CHROMATOGRAPHY**

**Method 314.0** was developed for the determination of perchlorate in drinking water by **ion chromatography coupled with suppressed conductivity detection**.

**Method 314.1** employs the use of **inline column concentration/matrix elimination ion chromatography coupled with suppressed conductivity detection** for the determination of perchlorate in drinking water.

**Method 332.0** was developed for the determination of low level perchlorate in drinking water by **ion chromatography coupled with suppressed conductivity and electrospray ionization mass spectroscopy**. To improve the accuracy and performance of this method, **<sup>18</sup>-Oxygen enriched perchlorate is utilized as an internal standard**.

**Method 314 Calibration**

1000 ug/mL in Water

Sodium Perchlorate (NaClO<sub>4</sub><sup>-</sup>)

**Part # 57001 \$35/100 mL**

**Part # 54164 \$150/500 mL**

100 ug/mL in Water

Sodium Perchlorate (NaClO<sub>4</sub><sup>-</sup>)

**Part # 59344 \$35/100 mL**

**Laboratory Synthetic  
Sample Matrix Blank (pH 8.6)**

1000 ug/mL in Water

(1) Bicarbonate (HCO<sub>3</sub><sup>-</sup>)

(2) Chloride (Cl<sup>-</sup>)

(3) Sulfate (SO<sub>4</sub><sup>2-</sup>)

**Part # 54205 \$50/100 mL**

**Laboratory Synthetic  
Sample Matrix Blank (pH 10)**

25 mg/mL in Water

(1) Carbonate (CO<sub>3</sub><sup>2-</sup>)

(2) Chloride (Cl<sup>-</sup>)

(3) Sulfate (SO<sub>4</sub><sup>2-</sup>)

**Part # 59206 \$60/100 mL**

**Method 332.0 Calibration**

1000 ug/mL in Water

Sodium Perchlorate (NaClO<sub>4</sub><sup>-</sup>)

**Part # 57001 \$35/100 mL**

**Part # 54164 \$150/500 mL**

100 ug/mL in Water.

Sodium Perchlorate (NaClO<sub>4</sub><sup>-</sup>)

**Part # 59344 \$35/100 mL**

**Method 332.0 Calibration  
Internal Standard**

1 ug/mL in Water

Sodium Perchlorate(<sup>18</sup>O) -(NaCl <sup>18</sup>O<sub>4</sub><sup>-</sup>)

**Part # 54022 \$30/ 1 mL**

**Laboratory Synthetic  
Sample Matrix Blank (pH 10)**

25 mg/mL in Water

(1) Carbonate (CO<sub>3</sub><sup>2-</sup>)

(2) Chloride (Cl<sup>-</sup>)

(3) Sulfate (SO<sub>4</sub><sup>2-</sup>)

**Part # 59206 \$60/100 mL**

**DETERMINATION OF NITRATE-NITRITE NITROGEN  
BY AUTOMATED COLORIMETRY**

METHOD

**353.2**

This method covers the determination of nitrite singly, or nitrite and nitrate combined in drinking, ground, and surface waters, and domestic and industrial wastes. Stability of the solution must be closely monitored. Detection is by colorimetry. Typical stability is only 1 month. Store the materials at 4°C.

**EPA Method 353.2 Anions Mix**

1000 ug/mL in Water.

- (1) Nitrate as N
- (2) Nitrite as N


**Part # 52120      \$25/100 mL****Part # 53120      \$100/500 mL****EPA Method 353.2 Anions Mix**

Varied ug/mL in Water.

- (1) Nitrate as N      1000
- (2) Nitrite as N      500

**Part # 54162      \$25/100 mL****Part # 54521      \$100/500 mL**

**Method:** E300D.M. Column: ASAHIPACK ODP50 (5 $\mu$ m X 250 X 4.0mm). Flow Rate= 1.0mL/min.. Column Temp.= 40°C. Isocratic Analysis using Anion Mobile Phase. Detector: PDA (DAD1, Sig=360.20 REF=266, 10).  
Analyst: Pedro Rentas.



METHOD

**Misc.**

**MISCELLANEOUS  
ORGANIC ACIDS /  
WET CHEMICAL CALIBRATION STANDARDS**

**Acetic Acid (Acetate)**

1000 ug/mL in Water.

Part # 54116     \$35/100 mL  
Part # 54516     \$150/500 mL

**Formic Acid (Formate)**

1000 ug/mL in Water.

Part # 54117     \$35/100 mL  
Part # 54517     \$150/500 mL

**Oxalic Acid (Oxalate)**

1000 ug/mL in Water.

Part # 54196     \$35/100 mL  
Part # 54596     \$150/500 mL

**Organic Acids Mix**

1000 ug/mL in Water.

- (1) Acetic acid
- (2) Formic acid
- (3) Methoxyacetic acid
- (4) Oxalic acid

Part # 59294     \$350/500 mL

**Tetramethyl Ammonium Hydroxide**

1000 ug/mL in Water.

Part # 54118     \$35/100 mL  
Part # 54518     \$150/500 mL

**Phenolics (Total) 4AAP**

1000 ug/mL in Water

Part # 54136     \$30/100 mL

**Total Organic Halides (TOX)**

as 2,4,6-Trichlorophenol

Cl @ 100 ug/mL

Part # 54194 in Water     \$25/ 20 mL  
Part # 54243 in Methanol \$25/ 20 mL

**Total Solids**

100 mg Neat

Part # 54167     \$25/ 100 mg

**Non-Filterable Solids  
(TSS)**

100 mg Neat

Part # 54134     \$25/ 100 mg

**Total Dissolved Solids  
(TDS)**

1000 ug/mL in Water.

Part # 54125     \$25/100 mL



## WET CHEMICAL CALIBRATION MISCELLANEOUS EPA & STANDARD METHODS

METHOD

**Misc.**

The following section contains several Wet Chemical Reference Materials that can be used as quality control checks or to generate calibration curves. A few popular EPA/SW-846 and Standard Methods are depicted for easy reference. Absolute Standards, Inc. has also generated a completely searchable web-based index of EPA analytes and methods. Please see our web site: [www.AbsoluteStandards.com](http://www.AbsoluteStandards.com)

**Organic Nutrients**

TKN as N, Total Phosphorus as P

1000 ug/mL in Water

**Part # 54152    \$25/100 mL****Total Organic Carbon (TOC) & Sulphate**

1000 ug/mL in Water

**Part # 54148    \$25/100 mL****Total Kjeldahl Nitrogen from Glycine**

1000 ug/mL in Water

**Part # 54132    \$25/100 mL****Total Organic Carbon (TOC) from KHP**

1000 ug/mL in Water

**Part # 59179    \$25/100 mL****Total Phosphorus as P**

1000 ug/mL in Water

**Part # 54133    \$25/100 mL****Total Organic Carbon (TOC) from Sucrose**

1000 ug/mL in 1% HCl/Water.

**Part # 54126    \$25/100 mL****Inorganic Nutrients**NH<sub>4</sub> as N, NO<sub>3</sub> as N, PO<sub>4</sub> as P

1000 ug/mL in Water

**Part # 54153    \$25/100 mL****Calcium Hardness**

1000 ug/mL in 1% HCl

**Part # 54143    \$25/100 mL****UV 254 Absorbance/  
Dissolved Organic Carbon (DOC)  
as KHP**100 ug/mL as DOC in Water  
14.4 cm<sup>-1</sup> as Absorbance at 254 nm in Water**Part # 54166    \$25/100 mL****Total Hardness**

1000 ug/mL in Water

**Part # 54156    \$25/100 mL****Part # 54597    \$90/500 mL****Alkalinity as CaCO<sub>3</sub>**

1000 ug/mL in Water

**Part # 54142    \$25/100 mL****Part # 54242    \$75/500 mL**

METHOD

**Misc.**

**WET CHEMICAL CALIBRATION  
MISCELLANEOUS EPA & STANDARD METHODS**

**Conductance @ 25 °C**

1000 uMHOS/cm in Water

Part # 54131 \$25/100 mL

**Turbidity**

400 NTU in Water

Part # 54122 \$25/20 mL

**pH of 6**

in Water

Part # 54119 \$25/100 mL

**MBAS**

1000 ug/mL in Water

Part # 54160 \$25/20 mL

**Residual Free Chlorine**

1000 ug/mL in Water

Part # 54124 \$25/100 mL

**Color**

500 Color Units in 10% HCl/H<sub>2</sub>O

Potassium Chloroplatinate  
Cobaltous Chloride hexahydrate

Part # 54255 \$75/100 mL

**Silica (SiO<sub>2</sub>)**

1000 ug/mL in Water / tr. NaOH

Part # 54159 \$25/100 mL

Part # 54169 \$75/500 mL

**Simple Cyanide**

**as Potassium Cyanide**

1000 ug/mL in Water/NaOH

Part # 59017 \$25/100 mL

**Total Cyanide**

**as Potassium Ferricyanide**

1000 ug/mL in Water/NaOH

Part # 54150 \$25/100 mL

**Simple & Total Cyanide**

1000 ug/mL in Water/NaOH

Part # 54165 \$25/100 mL

## PRODUCT NUMBER CROSS REFERENCE INDEX

## INDEX

Part#	Page#	Part #	Page#	Part#	Page#
52000	16	52073	46	52142	58
52001	16	52074	47	52143	58
52002	16	52075	47	52144	60
52003	16	52076	47	52145	60
52004	16	52079	49	52146	60
52005	16	52081	49	52147	60
52006	36	52084	50	52148	37
52007	36	52085	50	52149	19
52008	36	52086	51	52150	19
52009	36	52087	51	52151	19
52017	20	52088	51	52152	19
52018	20	52089	51	52153	19
52022	21	52090	51	52154	20
52023	21	52091	62	52155	23
52024	22, 30	52092	62	52156	23
52025	22	52093	62	52157	24
52028	22	52094	62	52158	24
52031	22	52095	63	52159	24
52038	39	52096	63	52160	25
52039	39	52097	63	52161	25
52041	39	52098	63	52162	26
52042	38	52100	54	52163	28, 31
52043	38	52101	54	52164	28, 31
52044	38	52102	54	52166	29
52045	38	52105	68	52167	30
52046	38	52106	68	52168	32
52047	38	52107	66	52169	32
52050	39	52114	55	52170	33
52056	41	52115	55	52171	33
52057	41	52116	55	52172	34
52058	42	52117	55	52173	34
52059	42	52118	66	52174	34
52060	42	52119	66, 68	52175	34
52061	43	52120	70	52176	35
52063	44	52122	52, 60	52177	35
52064	45	52123	52	52178	35
52065	45	52133	56	52250	15
52066	45	52134	56	52251	15
52067	45	52135	56	52252	57
52068	45	52136	56	52253	14
52069	46	52137	56	52254	14
52070	46	52139	58	52255	14
52071	46	52140	58	52256	14
52072	46	52141	58	52302	54

**INDEX****PRODUCT NUMBER  
CROSS REFERENCE INDEX**

<b>Part#</b>	<b>Page#</b>	<b>Part #</b>	<b>Page#</b>	<b>Part#</b>	<b>Page#</b>
52309	16	53050	39	53115	55
52310	16	53056	41	53116	55
52311	16, 57	53057	41	53117	55
52312	16	53058	42	53118	66
52313	16	53059	42	53119	66, 68
52314	16	53060	42	53120	70
52315	16	53061	43	53122	52, 60
52316	16	53063	44	53123	52
52319	52	53064	45	53133	56
52320	66	53065	45	53134	56
52331	40	53066	45	53135	56
52332	52	53067	45	53136	56
52333	53	53068	45	53137	56
52334	53	53069	46	53139	58
52335	41	53070	46	53140	58
52336	41	53071	46	53141	58
52337	41	53072	46	53142	58
53000	16	53073	46	53143	58
53001	16	53074	47	53144	60
53002	16	53075	47	53145	60
53003	16	53076	47	53146	60
53004	16	53079	49	53147	60
53005	16	53081	49	53148	37
53006	36	53084	50	53149	19
53007	36	53085	50	53150	19
53008	36	53086	51	53151	19
53009	36	53087	51	53152	19
53017	20	53088	51	53153	19
53018	20	53089	51	53154	20
53022	21	53090	51	53155	23
53023	21	53091	62	53156	23
53024	22, 30	53092	62	53157	24
53025	22	53093	62	53158	24
53028	22	53094	62	53159	24
53031	22	53095	63	53160	25
53038	39	53096	63	53161	25
53039	39	53097	63	53162	26
53041	39	53098	63	53163	28, 31
53042	38	53100	54	53164	28, 31
53043	38	53101	54	53165	29
53044	38	53105	68	53166	29
53045	38	53106	68	53167	30
53046	38	53107	66	53168	32
53047	38	53114	55	53169	32


## PRODUCT NUMBER CROSS REFERENCE INDEX

## INDEX


Part#	Page#	Part #	Page#	Part#	Page#	Part#	Page#
53170	33	54133	73	54596	72	59203	69
53171	33	54134	72	54597	73	59206	70
53172	34	54135	61	57001	70	59219	40
53173	34	54136	72	57003	69	59249	40
53174	34	54139	65	57011	69	59294	72
53175	34	54142	73	57012	69	59322	69
53176	35	54143	73	57019	69	59344	70
53177	35	54148	73	57020	69	59370	29
53178	35	54149	69	57038	69	59388	53
53309	16	54150	74	57056	69	91958	61
53310	16	54152	73	57146	16		
53311	16, 57	54153	73	57250	58		
53312	16	54154	69	57280	16		
53313	16	54155	65	57315	38		
53314	16	54156	73	57351	46		
53315	16	54157	69	57380	45, 46		
53316	16	54158	69	57642	43		
53319	52	54159	74	57651	45		
53320	66	54160	74	57847	60		
53331	40	54161	59	57880	56		
53332	52	54162	70	57911	39		
53333	53	54163	65	58146	16		
53334	53	54164	70	58250	58		
53335	41	54165	74	58280	16		
53336	41	54166	73	58351	46		
53337	41	54167	72	58380	45, 46		
54000	65	54169	74	58642	43		
54001	69	54172	59	58651	45		
54022	70	54175	59	58847	60		
54112	69	54176	59	58880	56		
54113	69	54194	72	58911	39		
54116	72	54196	72	59010	67		
54117	72	54205	70	59011	67		
54118	72	54242	73	59017	74		
54119	74	54243	72	59030	49		
54122	74	54255	74	59045	64		
54124	74	54500	65	59046	64		
54125	72	54516	72	59076	66		
54126	73	54517	72	59095	67		
54127	65	54518	72	59096	67		
54128	65	54521	70	59110	64		
54129	65	54527	65	59153	64		
54131	74	54528	65	59176	66		
54132	73	54529	65	59179	73		

**NOTES**

**NOTES**



# Absolute Standards, Inc.



**Custom Standard Quotation Request Form**  
**Fax:** (800) 410-2577, Technical Service Dept, Absolute Standards, Inc.  
**Email:** customerserv@absolutestandards.com

Page \_\_\_\_\_ of \_\_\_\_\_

Date: \_\_\_\_ / \_\_\_\_ / \_\_\_\_

**From:**

Company Contact: \_\_\_\_\_

Company: \_\_\_\_\_ Account#: \_\_\_\_\_

Company Address: \_\_\_\_\_

Telephone: \_\_\_\_\_ Fax: \_\_\_\_\_

Email: \_\_\_\_\_

**Product Description:** \_\_\_\_\_

**Solvent/ Matrix:** \_\_\_\_\_

**Volume Size:**  1 x 100 mL  1 x 500 mL  Other: \_\_\_\_\_

Qty: \_\_\_\_\_ Qty: \_\_\_\_\_ Qty: \_\_\_\_\_

Date Required: \_\_\_\_ / \_\_\_\_ / \_\_\_\_ Analysis requested? (additional charge)  Yes  No

<u>Component</u>	<u>CAS# (optional)</u>	<u>Concentration (ug/mL)</u>
1. _____	_____	_____
2. _____	_____	_____
3. _____	_____	_____
4. _____	_____	_____
5. _____	_____	_____
6. _____	_____	_____
7. _____	_____	_____
8. _____	_____	_____
9. _____	_____	_____
10. _____	_____	_____
11. _____	_____	_____
12. _____	_____	_____
13. _____	_____	_____
14. _____	_____	_____
15. _____	_____	_____
16. _____	_____	_____
17. _____	_____	_____
18. _____	_____	_____
19. _____	_____	_____
20. _____	_____	_____

Comments: \_\_\_\_\_